



PROFICIENCY TESTING REPORT
ISHTM-AIIMS EXTERNAL QUALITY ASSURANCE PROGRAMME
 NABL accredited program as per ISO/IEC 17043:2010 standard
 Organized By Department of Hematology, AIIMS, New Delhi-110029



Duration of stability testing - minimum upto 8 days at ambient temp. after dispatch of specimens

EQAP CODE No. : 5425

Distribution No.: 160-N

Month/Year: July/2023

Instrument ID: ERBA H560(K1104B2211056)

Name & Contact No. of PT Co-ordinator: Dr. Manoranjan Mahapatra (Prof. & Head), Hematology, AIIMS, Delhi,
 Tel: 9013085730 , E-Mail : accuracy2000@gmail.com

Date of issue & status of the report: 05-10-2023[Final].

CBC and Retic Assessment

Test Parameters	S.No.	Among Lab (Accuracy Testing)						Within Lab (Precision Testing)			
		Your Result 1	Your Result 2	Your Results Sum of 2 Value	Consensus result sum of 2 values (Assigned Value)	Uncertainty of Assigned Values	Z Score	Yours Results Diff. of 2 Values	Consensus Result Diff. of 2 values (Assigned Value)	Uncertainty of Assigned Values	Z Score
WBC x10 ³ /µl	1	3.99	3.87	7.86	9.3	0.036	-1.49	0.12	0.1	0.007	0.18
RBC x10 ⁶ /µl	1	4.21	4.16	8.37	8.35	0.011	0.07	0.05	0.04	0.002	0.27
Hb g/dl	1	10.5	10.4	20.9	21.5	0.026	-0.98	0.1	0.1	0.007	0.00
HCT%	1	33.3	32.7	66	68.4	0.163	-0.50	0.6	0.3	0.019	0.81
MCV-fl	1	79.1	78.8	157.9	165.1	0.312	-0.77	0.3	0.2	0.017	0.34
MCH-Pg	1	24.9	24.8	49.7	51.7	0.071	-1.12	0.1	0.2	0.014	-0.45
MCHC-g/dl	1	31.7	31.4	63.1	62.7	0.148	0.10	0.3	0.3	0.017	0.00
Plt. x10 ³ /µl	1	144	128	272	304	1.864	-0.66	16	6	0.361	1.69
Retic %	2	3.1	2.9	6	6.85	0.131	-0.22	0.2	0.4	0.022	-0.90

P.S . Assesment

YOUR REPORT			CONSENSUS REPORT		
DLC%	3	Nrbcs=01 , Poly=24 L=72, E=01, Mono/Promono=02 , B1= P.M.=, Mye=, Meta=, Other=MANY SMUDGE CELLS & FEW ATYPICAL LYMPHOID CELLSSEEN	Lymp: 68-79, Poly: 15-26, mono: 1-3,Eosino: 1-2, nRBC/blast/Myelo/Meta/: 0-5		
RBC Morphology	3	NORMOCYTIC NORMOCHROMIC TO MILD HYPOCHROMIC.	Predominantly: Normocytic/Normochromic; Moderate: Anisocytosis, Microcytic, Hypochromia; Mild: Poikilocytosis.		
Diagnosis	3	CHRONIC LYMPHOPROLIFERATIVE DISORDER	Chronic Lymphoproliferative Disorder/CLL		

COMBINED DATA VALUES OF TOTAL PARTICIPANTS

Test parameters	S.No.	Total participants covered in the current dist. 160--N	Total No. responded	% of Labs with Z Score 0-2		% of Labs with Z Score 2-3		% of Labs with Z Score >3	
				Among labs	Within lab	Among labs	Within lab	Among labs	Within lab
WBC x10³/µl	1	343	340	84.12	85.29	6.18	8.24	9.7	6.47
RBC x10⁶/µl	1	343	343	84.55	90.96	7	3.79	8.45	5.25
Hb g/dl	1	343	343	86.88	90.09	5.83	4.66	7.29	5.25
HCT%	1	343	341	91.5	87.1	4.4	5.57	4.1	7.33
MCV-fl	1	343	341	93.26	87.1	3.81	3.23	2.93	9.67
MCH-Pg	1	343	340	87.06	94.12	7.94	2.35	5	3.53
MCHC-g/dl	1	343	341	92.38	86.51	4.69	5.28	2.93	8.21
Plt. x10³/µl	1	343	341	88.27	89.74	7.62	5.57	4.11	4.69
ReticCount%	2	343	266	90.98	80.08	6.77	14.29	2.25	5.63
PS Assessment	3	343	256	Satisfactory :95.03%, Borderline Sat. :1.75%, Unsatisfactory :3.22%					

***Comments:**

1). Among Lab (EQA) : Results acceptable.

2). Within Lab (IQA) : Precision acceptable.

Note-1: EQA (External Quality Assurance) : Your Performance among various of participating labs in PT, to determine the accuracy of your results.

IQA (Internal Quality Assurance) : Your Performance of comparison of two consecutive measurement values within your lab to test the precision of your autoanalyzer.

Note-2: Z score among & within lab were calculated, as per to ISO/IEC 13528:2015 standard. Z score among lab (EQA)= (Your Result Sum of two values - Consensus Result sum of two values)/(Normalised IQR)

Z score within lab (IQA)= (Your Result Difference of two values - Consensus Result difference of two values)/(Normalised IQR)

IQR = Quartile 3 - Quartile 1 of participant data, Normalised IQR = 0.7413 x IQR

Note-3: Z score 0 to ± 2 : Acceptable, Z score ± 2 to ± 3 :Warning Signal, Z score $> \pm 3$: Unacceptable [As per ISO/IEC 13528:2015 standard]

Note-4: Z score value between "0 to ± 2 " are texted in green colour. Z score value between " ± 2 to ± 3 " are texted in orange colour. Z score value $> \pm 3$ are texted in red colour.

Note-5: Homogeneity and stability testing of PT sample were done as per ISO 13528:2015 standard. To pass homogeneity test, between sample SD (Ss) should be smaller than the check value (0.3*SDPA). To pass the stability test, average difference in measurement values of first and last day sample ($\bar{x}-\bar{y}$) should be smaller than the check value (0.3*SDPA).

Note-6: ISHTM-AIIMS-EQAP does not subcontract any task of its scheme

Note-7: Participants are free to use methods/analyzer of their own choice.

Note-8: Proficiency testing (PT) samples are sent quarterly to each participant.

Note-9: All the necessary details regarding design and implementation of PT, are provided in the instruction sheet as well as on programme's website www.ishtmaiimseqap.com.

Note 10: Reports are kept confidential.

Report authorized by,



Dr. Manoranjan Mahapatra (Prof. & Head)

PT Co-ordinator: ISHTM-AIIMS-EQAP

Department of Hematology, AIIMS, New Delhi

-----End Of Report-----