



**PROFICIENCY TESTING REPORT**  
**ISHTM-AIIMS EXTERNAL QUALITY ASSURANCE PROGRAMME**  
 NABL accredited program as per ISO/IEC 17043:2010 standard  
 Organized By Department of Hematology, AIIMS, New Delhi-110029



*Duration of stability testing - minimum upto 8 days at ambient temp. after dispatch of specimens*

EQAP CODE No. : 3400

Distribution No.: 160-I

Month/Year: June/2023

Instrument ID: Mindray BC 5000 (SS-75004270)

Name & Contact No. of PT Co-ordinator: Dr. Manoranjan Mahapatra ( Prof. & Head), Hematology, AIIMS, Delhi,  
 Tel: 9013085730 , E-Mail : accuracy2000@gmail.com

Date of issue &amp; status of the report: 08-08-2023[Final].

### CBC and Retic Assessment

Test Parameters	S.No.	Among Lab (Accuracy Testing)						Within Lab (Precision Testing)			
		Your Result 1	Your Result 2	Your Results Sum of 2 Value	Consensus result sum of 2 values (Assigned Value)	Uncertainty of Assigned Values	Z Score	Yours Results Diff. of 2 Values	Consensus Result Diff. of 2 values (Assigned Value)	Uncertainty of Assigned Values	Z Score
WBC x10 <sup>3</sup> /µl	1	5.71	5.61	11.32	12	0.089	-0.36	0.1	0.1	0.013	0.00
RBC x10 <sup>6</sup> /µl	1	4.76	4.75	9.51	9.61	0.014	-0.44	0.01	0.04	0.004	-0.67
Hb g/dl	1	15.7	15.6	31.3	30.3	0.054	1.08	0.1	0.1	0.011	0.00
HCT%	1	54.5	53.6	108.1	92.3	0.320	2.52	0.9	0.4	0.041	1.12
MCV-fl	1	114.4	113.2	227.6	193	0.544	3.25	1.2	0.3	0.031	2.43
MCH-Pg	1	32.9	32.9	65.8	63	0.106	1.43	0	0.3	0.023	-1.35
MCHC-g/dl	1	29.1	28.7	57.8	64.6	0.229	-1.52	0.4	0.3	0.028	0.34
Plt. x10 <sup>3</sup> /µl	1	142	131	273	305	2.716	-0.71	11	5	0.459	1.08
Retic %	2	2	1.8	3.8	5.4	0.169	-0.50	0.2	0.3	0.032	-0.27

### P.S . Assesment

YOUR REPORT		CONSENSUS REPORT
<b>DLC%</b>	3	Nrbcs=2 , Poly=20 L=08, E=00, Mono/Promono=04 , B1=36 P.M.=15, Mye=10, Meta=03, Other=
<b>RBC Morphology</b>	3	Predominantly: Normocytic/Normochromic; Moderate: Microcytosis, Hypochromia; Mild: Anisocytosis
<b>Diagnosis</b>	3	Acute Myeloid Leukemia (AML)

**COMBINED DATA VALUES OF TOTAL PARTICIPANTS**

Test parameters	S.No.	Total participants covered in the current dist. 160--I	Total No. responded	% of Labs with Z Score 0-2		% of Labs with Z Score 2-3		% of Labs with Z Score >3	
				Among labs	Within lab	Among labs	Within lab	Among labs	Within lab
<b>WBC x10<sup>3</sup>/µl</b>	1	145	143	86.71	86.71	4.2	4.9	9.09	8.39
<b>RBC x10<sup>6</sup>/µl</b>	1	145	145	84.14	88.97	7.59	2.07	8.27	8.96
<b>Hb g/dl</b>	1	145	145	85.52	86.9	6.21	6.9	8.27	6.2
<b>HCT%</b>	1	145	143	93.01	90.21	4.9	5.59	2.09	4.2
<b>MCV-fl</b>	1	145	143	93.71	88.11	4.2	7.69	2.09	4.2
<b>MCH-Pg</b>	1	145	143	87.41	93.01	5.59	2.8	7	4.19
<b>MCHC-g/dl</b>	1	145	143	93.71	90.21	4.2	2.8	2.09	6.99
<b>Plt. x10<sup>3</sup>/µl</b>	1	145	143	90.21	93.01	8.39	1.4	1.4	5.59
<b>ReticCount%</b>	2	145	134	92.54	95.52	5.97	0.00	1.49	4.48
<b>PS Assessment</b>	3	145	129	Satisfactory :91.05%, Borderline Sat. :2.06%, Unsatisfactory :6.89%					

**\*Comments:**

- 1). Among Lab (EQA) : CBC result for MCV unacceptable, may be due to random/human error**
- 2). Within Lab (IQA) : Precision acceptable.**

**Note-1: EQA** (External Quality Assurance) : Your Performance among various of participating labs in PT, to determine the accuracy of your results.

**IQA** ( Internal Quality Assurance) : Your Performance of comparison of two consecutive measurement values within your lab to test the precision of your autoanalyzer.

**Note-2:** Z score among & within lab were calculated, as per to ISO/IEC 13528:2015 standard. Z score among lab (EQA)= (Your Result Sum of two values - Consensus Result sum of two values)/(Normalised IQR)

Z score within lab (IQA)= (Your Result Difference of two values - Consensus Result difference of two values)/(Normalised IQR)

IQR = Quartile 3 - Quartile 1 of participant data, Normalised IQR = 0.7413 x IQR

**Note-3:** Z score 0 to  $\pm 2$ : Acceptable, Z score  $\pm 2$  to  $\pm 3$  :Warning Signal, Z score  $> \pm 3$  : Unacceptable [As per ISO/IEC 13528:2015 standard]

**Note-4:** Z score value between "0 to  $\pm 2$ " are texted in green colour. Z score value between " $\pm 2$  to  $\pm 3$ " are texted in orange colour. Z score value  $> \pm 3$  are texted in red colour.

**Note-5:** Homogeneity and stability testing of PT sample were done as per ISO 13528:2015 standard. To pass homogeneity test, between sample SD (Ss) should be smaller than the check value (0.3\*SDPA). To pass the stability test, average difference in measurement values of first and last day sample ( $\bar{x}-\bar{y}$ ) should be smaller than the check value (0.3\*SDPA).

**Note-6:** ISHTM-AIIMS-EQAP does not subcontract any task of its scheme

**Note-7:** Participants are free to use methods/analyzer of their own choice.

**Note-8:** Proficiency testing (PT ) samples are sent quarterly to each participant.

**Note-9:** All the necessary details regarding design and implementation of PT, are provided in the instruction sheet as well as on programme's website [www.ishtmaiimseqap.com](http://www.ishtmaiimseqap.com).

**Note 10:** Reports are kept confidential.

Report authorized by,



Dr. Manoranjan Mahapatra ( Prof. & Head)

PT Co-ordinator: ISHTM-AIIMS-EQAP

Department of Hematology, AIIMS, New Delhi

-----End Of Report-----