



# RML- Quality Assurance Program (RML-QAP)



## Urine Routine Analysis

### ALL METHOD REPORT

Lab Code: 3064

Cycle - 12/2023  
Round No -09

Date: 05/10/2023

Parameters	Units	No.of. Participants	Consensus		Consensus Result	Your Value	Remark
			Number	%			
Specific Gravity	-	352	316	90%	1.010-1.020	1.020	Within Consensus
pH	-	352	315	90%	6.0	6.0	Within Consensus
Leukocytes	/µl	345	337	98%	Negative	Negative	Within Consensus
Nitrite	-	342	337	99%	Negative	Negative	Within Consensus
Protein	mg/dl	351	345	98%	Negative	Negative	Within Consensus
Glucose	mg/dl	351	349	99%	Negative	Negative	Within Consensus
Ketones	-	351	348	99%	Negative	Negative	Within Consensus
Urobilinogen	mg/dl	349	344	99%	Negative	Negative	Within Consensus
Bilirubin	mg/dl	349	345	99%	Negative	*2+	Outside Consensus
Blood (Hemoglobin)	/µl	349	341	98%	Negative	Negative	Within Consensus

Legends	≥ 60% Consensus	(*) Outside Consensus	(@) No Consensus	(.) Not Reported	(#) Late Result Submission
---------	-----------------	-----------------------	------------------	------------------	----------------------------

Chief Coordinator

Dr. Sanjay Mehrotra

Checked By:

Prepared By: SM

Programme Director

Dr. Bandana Mehrotra

Page 1 of 4

Doc. No.: ASS/FR/19/R 01/Dt.: 05.01.2022



Address: B-171, Nirala Nagar, Lucknow-226020. Ph.: 4034100-130 (30 Lines), 4077180, 2788444 Fax: (0522)2788555  
Email: rmlresearchfoundation@gmail.com Website: www.rmlpathology.com

Continuous Efforts And Execution Leads To Quality Excellence