

06-NOV-2023

### CERTIFICATE OF CALIBRATION

Name and Address of Customer : Checkpoint Healthcare Pvt. Ltd.

Location : P-41, Pandav Nagar, Mayur Vihar Phase-1  
Delhi-110091

Model : Automated Hematology Analyzer Sysmex XQ -320

Make : Sysmex, Japan

Serial No : 12016

Calibration Date : 16-09-2023

Next Calibration Date : 15-09-2024

Material Used : SCS-1000 ( LOT NO. 32480525 )

Expiry : 08-10-2023

**Note: This is to certify that the above mentioned instrument installed in the lab has been calibrated using the above mentioned calibrator satisfactorily. The following procedure were done and passed ok.**

S.NO	PROCEDURE	PASS
1.	Background Check	OK
2.	Precision	OK
3.	Calibrator Run	OK
4.	Tri Level QC run	OK

For Meridian Life Care

