



ASIAN TECHNOLOGY

(A HOUSE OF CALIBRATION)



CALIBRATION CERTIFICATE

Customer name And Address	M/S: Checkpoint Healthcare Pvt. Ltd. P-41, Pandav Nagar Mayur Vihar Phase-1, Delhi-110091	Service request No. & date	P/03/07/09/23&07/09/2023
		ULR No.	CC223923000004987F
		Cert. No.	AT/23000004987
		Date of Receipt of DUC	07/09/2023
		Date of calibration	07/09/2023
		Date of issue	08/09/2023
		Suggested due date	06/09/2024

Instrument Details			
Instrument name	Micro Pipette	Sr. No.	AB154235
Make	Erba Biohit	Location	----
Range / Size	100 to 1000 µl	Accuracy	-----
Least Count	5 µl	Visual Inspection	OK
I.D. No.	MP-01	Calibration Performed at	Lab

Detail of reference standards & Major equipments used			
Equipment Name	A set of weight box	Digital Weighing Balance	
Make	WEIGHTRONICS	AND	
Model / SR No.	-----	GH-252/ 78002	
Certificate No.	TYCON/WB/12/274	TYCON/WB/02/23/327	
Calibration Validity	25/12/2023	21/10/2024	
Calibration by	Tycon Engineering	Tycon Engineering	

Environmental Condition	Temperature	(23± 3) °C	Calibration Reference	IS:4162 (1985)
	Relative Humidity	(50±20) %	Work Instruction	AT/WI-20M

Calibration Results			
Serial No.	Nominal Capacity in (µl)	Measured Capacity in (µl)	Uncertainty At 95% C.L. (coverage factor K=2)
01.	100	100.13	±0.46 µl
02.	300	300.39	
03.	500	499.76	
04.	800	798.84	
05.	1000	999.92	

Remarks:

- ❖ (1) Standard equipment use for calibration are traceable to national/ international standards.
- ❖ (2) The reported expanded uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by coverage factor k = 2 such that the coverage probability corresponds to approximately 95%. (3) The above results are valid at the time of and under the stated conditions measurement. (4) This certificate is refers only to the particular item submitted for calibration. (5) Next calibration due date given as requested by the customer.

Calibrated By
(Calibration Engg./TM)
(MAHIPAL)
Form No.- QF-47

Approved By
(QM/TM)
(NENRAJYAGI)
Page No. 1 of 1



---End of Report---

LAB.: F.C.A.-3614, S.G.M. NAGAR, NEAR RAJA CHOWK, N.I.T. FARIDABAD-121 001 (HARYANA)
Customer Care : 9654550258, E-mail : asiantech2011@gmail.com