

## **ASIAN TECHNOLOGY**

(A HOUSE OF CALIBRATION)



## **CALIBRATION CERTIFICATE**

Customer name And Address	M/S: Checkpoint Healthcare Pvt. Ltd. P-41, Pandav Nagar Mayur Vihar Phase-1, Delhi-110091	Service request No. & date ULR No. Cert. No. Date of Receipt of DUC Date of calibration Date of issue	P/03/07/09/23&07/09/2023 CC223923000004992F AT/23000004992 07/09/2023 07/09/2023 08/09/2023
	*	Suggested due date	06/09/2024

Instrument Details			
Instrument name	Digital Thermometer	Location	Refrigerator Down
Make	may 1	Accuracy	
Range / Size	-50 to 80 °C	Visual Inspection	OK .
Least Count	0.1 °C	Calibration Performed at	Site
I.D. No. / Sr. no.	DTM-02 /		*   500 to 1

			140	
	Detail of reference standards & I	Major equipments used	· · · · · · · ·	Te 14.
Equipment Name	SPRT With Digital Indicator		1 4 5 5	
Make	Tempsans/Fluke ·			
·Model / SR No.	05640/AT/TH/12		-1	1
Certificate No.	CC288822000001839 F		1 , , ,	
Calibration Validity	27/10/2023	W 10 10 10 10 10 10 10 10 10 10 10 10 10	7	
Calibration by	Sigma	2	1 1 2	7.

	• 4			
E-vivonmental	Temperature	(25±4)°C	Calibration Reference	IS:6274 (1971)
Environmental Condition	Relative Humidity	(50±20) %	Work Instruction	AT/WI-01TH
Condition	Terative Training	(		

Calibration Results

Serial No.	Std. Reading in (°C)	Measured Reading in  (°C)	Uncertainty At 95% C.L. (coverage factor k=2)
01.	2.012	2.1	
02.	3.021	3.2	±0.3 °C
03.	5.023	5.3	10.5
04.	6.032	6.2	
05.	7.281	8.0	The second secon

## Remarks:

- (1) Standard equipment use for calibration are traceable to national/international standards. (2)The reported expended uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by coverage factor The reported expended uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by coverage factor k = 2 such that the coverage probability corresponds to approximately 95% (3)The above results are valid at the time of and under the stated conditions measurement. (4)This certificate is refers only to the particular item submitted for calibration. (5) Next calibration due date given as requested by the customer.

Calibrated By (Calibration Engg./TM) (DINSAXENA) Form No.- QF-47



Approyed By (QM//TM) (NEERATTYAGI) Page No. 1 of 1

---End of Report---

LAB.: F.C.A.-3614, S.G.M. NAGAR, NEAR RAJA CHOWK, N.I.T. FARIDABAD-121 001 (HARYANA)

Customer Care: 9654550258, F. mail: asigntech2011@gmail.com Customer Care: 9654550258, E-mail: asiantech2011@gmail.com