

Parashar Micro Measurement Pvt. Ltd.

B-59, SECTOR-64, NOIDA (U.P.) 201 307
PH.: 0120-4252360-65, Email : pmmpl.india@gmail.com
Website : www.pmmpl.in, www.parashar.info



Calibration Certificate

ULR - CC279523000068738F Page 1 of 1

Certificate Number : PMM/021223/11-1		Date of Request	02.12.2023
Calibrated For	M/s. Central Lab Clinical Pathology	Calibrated on	02.12.2023
	Jawahar Lal Nehru Medical College & Hospital	Valid Upto	01.12.2024
	Bhagalpur, Bihar.	Cert. Issue Date	04.12.2023

Description of Equipments

Nomenclature	Centrifuge	Accuracy/Class	Not Mentioned
Range	0 - 6000 rpm	Condition of UUC	Physically Ok
Resolution / Least Count	10 rpm	Serial No.	ZEBN-08199
Make & Model No.	Remi / Neya 4	Party ID Mark No.	-----
Material / Type	-----	Location	-----

Master Equipment / Standard Used

Sr.No.	Nomenclature	Make & Model	Serial No. / I/D Mark	Calibrated From	Certificate No.	Due Date of Calibration
1	Dig. Tachometer	Metravi	PMM/TM/09	PMMPL, Noida.	PMM/1418	04.12.2023

Standard used for calibration are traceable to Accredited lab, for Standard ISO/IEC : 17025 - 2017 or National Standards Through unbroken chain of calibration

Calibration Performed at	Site	Environmental	Temperature = 23 ± 5° C
Calibration Procedure No.	CP/MMV/S/03	Condition(s)	Humidity = 30 to 75 % RH
Reference Standard / Guideline	-----		

Mechanical Calibration


CALIBRATION RESULTS

Sr.No.	Nominal Value in UUC (in rpm)	Actual Observed Value in Std. (in rpm)	Error (in rpm)	± Uncertainty in Measurement (in rpm)
1	1000	1002	-2	51
2	2000	2003	-3	51
3	3000	3005	-5	51
4	4000	4007	-7	51
5	5000	5009	-9	51

UUC - Unit Under Calibration

Std. - Standard Instrument

Uncertainty of Measurement (at approx 95% Confidence Level with Coverage factor $k = 2$)

Calibrated By : 
Sudhir Kumar
Designation : Calibration Engineer



Approved By : 
Jai Prakash
Designation : Tech. Manager / Auth. Sign.

- Conditions :
1. This certificate refers only to the particular item submitted for calibration.
 2. The calibration result reported in this certificate are valid at the time of and under the stated conditions of measurement.
 3. This particular certificate can not be reproduced except in full, without prior permission of chief executive officer of the lab.