

Parashar Micro Measurement Pvt. Ltd.

B-59, SECTOR-64, NOIDA (U.P.) 201 307
PH.: 0120-4252360-65, Email : pmmpl.india@gmail.com
Website : www.pmmpl.in, www.parashar.info



The way to Technology
(20 Years of Excellence)



CC - 2795


Calibration Certificate

ULR - CC279523000068739F		Page 1 of 1				
Certificate Number : PMM/021223/11-2		Date of Request				
Calibrated For		Calibrated on				
M/s. Central Lab Clinical Pathology Jawahar Lal Nehru Medical College & Hospital Bhagalpur, Bihar.		02.12.2023				
		Valid Upto				
		01.12.2024				
		Cert. Issue Date				
		04.12.2023				
Description of Equipments						
Nomenclature	Centrifuge	Accuracy/Class	Not Mentioned			
Range	0 - 6000 rpm	Condition of UUC	Physically Ok			
Resolution / Least Count	10 rpm	Serial No.	ZEBN-08200			
Make & Model No.	Remi / Neya 4	Party ID Mark No.	-----			
Material / Type	-----	Location	-----			
Master Equipment / Standard Used						
Sr.No.	Nomenclature	Make & Model	Serial No. / I/D Mark	Calibrated From	Certificate No.	Due Date of Calibration
1	Dig. Tachometer	Metravi	PMM/TM/09	PMMPL, Noida.	PMM/1418	04.12.2023
Standard used for calibration are traceable to Accredited lab, for Standard ISO/IEC : 17025 - 2017 or National Standards Through unbroken chain of calibration						
Calibration Performed at		Site		Environmental		Temperature = 23 ± 5° C
Calibration Procedure No.		CP/MMV/S/03		Condition(s)		Humidity = 30 to 75 % RH
Reference Standard / Guideline		-----				
Mechanical Calibration						
CALIBRATION RESULTS						
Sr.No.	Nominal Value in UUC (in rpm)	Actual Observed Value in Std. (in rpm)		Error (in rpm)	± Uncertainty in Measurement (in rpm)	
1	1000	1001		-1	51	
2	2000	2002		-2	51	
3	3000	3004		-4	51	
4	4000	4006		-6	51	
5	5000	5008		-8	51	

UUC - Unit Under Calibration

Std. - Standard Instrument

Uncertainty of Measurement (at approx 95% Confidence Level with Coverage factor $k = 2$)

Calibrated By : 
Designation : Calibration Engineer



Approved By : 
Designation : Tech. Manager / Auth. Sign.

- Conditions : 1. This certificate refers only to the particular item submitted for calibration.
2. The calibration result reported in this certificate are valid at the time of and under the stated conditions of measurement.
3. This particular certificate can not be reproduced except in full, without prior permission of chief executive officer of the lab.