

10thSept2023

CalibrationCertificate

Name&Addressofthe customer	SUMANHOSPITAL,SULTANPUR
Make	Transasia
ProductName	ErbaEM200-FullyAutomated (BiochemistryAnalyzer)
SerialNo./InstrumentID	B200712

This is to certify that the cell counter installed at the above mentioned customer site has been calibrated with (Lot # 41000092, Expiry Date: 22/09/2023)

Sl. No.	Procedures	Remarks
1	Startup	PASS
2	Reproducibility	PASS
3	QCBio-Rad	PASS
4	Carryover	PASS
5	Calibration; Calibrator(Lot#410092)	PASS
6	Calibration Verification(c.f.a.s..Lot:4714920001)	PASS

Date of Calibration: 10/09/2023

Due date for Calibration: 09/09/2024

For Transasia Bio-Medicals Ltd.,


Mr. Gobinda Acharya

(Sr. Manager Service Operation)

