

# SIMCO CALIBRATION LABORATORY

(A Division of : Sharp Industrial Machinery Maintenance Co. Pvt. Ltd.)

10-3-74/27, Plot No.151, Street No:3, Teacher's Colony, East Marredpally, Secunderabad - 26. Telangana.

Tel: 040 2773 2341/342, 2773 1510, Mob: 77299 91231/33, 98480 46524,

E-mail : simco.hyd@gmail.com, Info@simco-india.com www.simcocalibrationlaboratory.com

**SIMCO**<sup>TM</sup>  
AN ISO 9001 : 2015  
CERTIFIED COMPANY



CC-2806

## CALIBRATION CERTIFICATE

In accordance with ISO / IEC-17025 ; 2017

F10-CC-03

Page : 1 of 2

<b>Certificate No. : SL2304TL0287-011</b>	<b>Issue Date : 04-05-2023</b>
<b>1. Customer Name &amp; Address:</b> M/s.State Reference Laboratory Of HIV Testing (APSACS), Department Of Microbiology, Kurnool Medical College, Kurnool, 518 002.	<b>ULR - C C 2 8 0 6 2 3 0 0 0 0 2 6 1 4 F</b>
	Reference Date : 29-04-2023 Calibration Date : 30-04-2023 Calibration Due Date : 29-04-2024

### 2. Details of Unit Under Calibration:

Description : Digital Thermo Hygrometer With Probe	Model No. : HTC-2
Make : HTC	Id No. : APSACS/ICTC/CHC/PKD/THM/01
Range : -50 to 70 °C / 10 to 99 %RH	
Resolution : 0.1 °C / 1 %RH	

### 3. Detail of Standard Instruments Used :

Instrument Used	SI / Id No	Valid up to	Certificate No.
Digital Thermo Hygrometer With Sensor	5180091/2027760 4	16-07-2023	30056478
Digital Temperature Indicator With SPRT Sensor	935-14-95 H	20-04-2024	TSC/22-23/18859-2

**4. Environmental Conditions:** Standard Temperature : (25±4) °C Relative Humidity : (50±20) % RH

**5. Calibration Procedure:** SOP-TL-01 / SOP-TL-04

**6. Thermal Calibration: Temperature & Relative Humidity**

### 7. Remarks:

- The instrument/equipment is in good condition and was calibrated at Lab.
- This certificate pertains only to the item calibrated.
- The calibration results reported in this certificate are valid at the time of and at the stated environmental conditions.
- The calibration interval is determined based on customer's requirements.
- The calibration is traceable to National standards as per traceability details given in the certificate.
- This calibration certificate shall not be reproduced in full, except with prior written approval of Managing Director, SIMCO Calibration Laboratory.
- This calibration certificate is meant for scientific and industrial purpose only.
- The NABL Symbol is used as per NABL guidelines in NABL-133.
- The Expanded Uncertainty is reported approximately at 95.45% confidence level with coverage factor  $k=2$ .

*Shingji*  
Calibrated by

*P.A. Anandam*  
Mrs. P.A. Anandam  
Technical Head  
Authorised Signatory

# SIMCO CALIBRATION LABORATORY

(A Division of : Sharp Industrial Machinery Maintenance Co. Pvt. Ltd.)

10-3-74/27, Plot No.151, Street No:3, Teacher's Colony, East Marredpally, Secunderabad - 26. Telangana.

Tel: 040 2773 2341/342, 2773 1510, Mob: 77299 91231/33, 98480 46524,

E-mail : simco.hyd@gmail.com, Info@simco-india.com www.simco Calibration Laboratory.com



CC-2806

## CALIBRATION CERTIFICATE

In accordance with ISO / IEC-17025 ; 2017

F10-CC-03

Page : 2 of 2

Certificate No. : SL2304TL0287-011

ULR - CC 2 8 0 6 2 3 0 0 0 0 2 6 1 4 F

### 8. Calibration Results:

#### a. Temperature @ 50% RH

S. No.	Standard Reading (°C)	UUC Reading (°C)	Error (°C)	Expanded Uncertainty in (±°C)
1	9.96	10.1	0.14	0.9
2	19.94	20.2	0.26	0.9
3	39.91	40.3	0.39	0.9
4	59.95	60.4	0.45	0.9

#### b. Relative Humidity @ 29°C

S. No.	Standard Reading (%RH)	UUC Reading (%RH)	Error (%RH)	Expanded Uncertainty in (± %RH)
1	19.68	20	0.32	1.2
2	34.49	35	0.51	1.2
3	60.37	61	0.63	1.2
4	95.04	96	0.96	1.2

#### c. Temperature (Probe)

S. No.	Standard Reading (°C)	UUC Reading (°C)	Error (°C)	Expanded Uncertainty in (±°C)
1	-30.075	-30.2	-0.125	0.4
2	0.007	0.1	0.093	0.4
3	10.055	10.2	0.145	0.4
4	40.081	40.3	0.239	0.4
5	60.053	60.4	0.347	0.4

*Shivaji*  
Calibrated by

*P. Anandam*  
Mrs. P.A. Anandam  
Technical Head  
Authorised Signatory