


CENTRIFUGE Calibration / Verification Record Sheet				
Location: ICTC-CHC-PATTIKONDA-KURNOOL			Range : Up to 3500RPM	
Equipment ID: APSACS/ICTC/CHC/PKD/CEF/01			SL.NO : 8907471-AP1Z-KNC-00196	
Calibration Procedure Reference: In-house verification of centrifuge using certified standard calibrated by Tachometer				
Date of Calibration :		08.12.2023	Due date of Calibration: 07.12.2024	
Tachometer ID : KMC/MB/HIVTL/SRL/TM/01			Calibration date : 19.04.2023	Due date of Calibration : 18.04.2024
Verification Done by: B.BHASKAR, Technical Officer, SRL, KMC,KURNOOL				
S.NO	STANDARD READING(RPM)	UUC READING (RPM)	ERROR (RPM)	
1.	499.3	500	0.7	
2.	998.7	1000	1.3	
3.	1997.6	2000	2.4	
4	2996.6	3000	3.4	
5.	3496.2	3500	3.8	
Accept/ Reject: Accepted since average difference is within +/- 100RPM				
Final Interpretation: Centrifuge is in good working condition				
Action Taken: - NA				
Remark: - NA				
				 TECHNICAL OFFICER