

CALIBRATION CERTIFICATE

Format No QSP-012/F01-00

<u>Name & Address of Customer:</u>		Certificate No.	: HTL/23-24/1/302-0001	
ICTC LAB AREA HOSPITAL Devarakonda, Nalgonda Dist, Telanagana.		Page No.	: 1 of 1	
		Place of Calibration	: Site	
Condition of UUC (At the time of receipt) Satisfactory	Equipment Receipt Date 23/11/2023	Date of Calibration 23/11/2023	Suggested Next Due 22/11/2024	Certificate Issue Date 25/11/2023

Details of Unit Under Calibration:

Instrument Name : Digital Timer
Make : SPRINTER
Model : -- **Range** : 0 to 9:59:59 Hrs
Id. No. : ICTC/TIM/01 **Resolution** : 1 Sec

Details of Reference Standard:

Nomenclature of Standard	Lab Id. No. / Sr. No.	Certificate No.	Calibration Due
Stop Watch	HTL/ET/002	MKT/23-24/725-01	13/11/2024


Environmental Condition: Temperature: 25 ± 4 °C, Relative Humidity: **30 to 75 %RH.**

Results of Calibration:

Sr. No.	Set Reading	STD Reading	UUC Reading	Error
	in Min	Min:Sec:1/100 Sec	Hr:Min:Sec	Min:Sec:1/100 Sec (C-B)
	A	B	C	
1	10	10:00:15	0:10:00	0:00:15
2	20	20:00:06	0:20:00	0:00:06
3	30	30:00:11	0:30:00	0:00:11
4	45	45:00:23	0:45:00	0:00:23
5	60	60:00:19	0:60:00	0:00:19

Remarks:

- a) The Standard used for calibration traceable to National/International standards through unbroken chain of Accredited Laboratories.
- b) Certificate shall not be reproduced, except in full, without the written approval of Horizon Techno Labs.
- c) Any corrections / modifications in the certificate should be brought to our knowledge within 30days from the date of issue of the certificate.
- d) This certificate refers to the values obtained at the time of calibration and under the stated conditions.
- e) The results stated in this certificate relate only to the item stated above.


 Calibrated By
Jaya Chandra
 Calibration Engineer


 Certificate Approved By
Shahbaz S F
 Head Technical Operations

End of Certificate

