



# UNIQUE CALIBRATION & CONSULTANT SERVICES

AN ISO 9001:2015 CERTIFIED CALIBRATION LABORATORY



CC-3287

## CALIBRATION CERTIFICATE

Format No.: UCCS-QF-7.8.1	Certificate No.: UCCSCC20230213-001	ULR No.: CC328723000002497F			
Group: Volume	Discipline: Mechanical				
Customer's Name & Address M/s. Dr. Bharti Path Lab Centre 1, Opp. PNB Bank, Mohna Road, Ballabhgarh	Page No.: 1 of 1	Calibration Performed: At Lab			
	Date of Receipt: 13-02-2023				
	Service Request No.: UCCSSRN20230213-001				
	Certificate Issue Date: 15-02-2023				
	Date of Calibration: 14-02-2023				
Next Due Date (Suggested by Customer): 13-02-2024					
Equipment Description (UUC) & Identification					
Name of the Instrument: Micro Pipette	Make: Erba				
	Model: -----				
Range: 5 µl to 50 µl	Sr. No.: -----				
Least Count: 0.5 µl	I.D. No.: MP-01				
Condition of Instrument: Ok	Location: Blood Bank				
Master Equipment Used					
Sr. No.	Instrument Name	Range & I.D. No.	Calibration Agency	Calibration Certificate No.	Valid Up to
1	Weight Set (E1 Class)	1 mg to 200 g & UCCS/CAL/048	Weightronics	WMCL/E/2022-11/2116	07-11-2025
2	Digital Weighing Balance	0 to 220 g & UCCS/CAL/005	Parashar Micro Measurement Pvt. Ltd.	PMM/101122/14-1	09-11-2023
Environmental Conditions			Calibration Procedure	Calibration Reference Standard	
Temperature: 20 ± 2 °C, Humidity: 50 ± 10 %RH Atmospheric Pressure : 14.2 psi			UCCS 03 SOP 018	ISO 8655	

### CALIBRATION RESULTS

Sr. No.	Nominal Capacity (µl) at 20 °C	Actual Volume (µl) at 20 °C	Error (µl)	Expanded Uncertainty (±) µl
1	10	10.39	-0.39	0.46
2	20	20.63	-0.63	0.46
3	50	50.88	-0.88	0.46

\*\*\*End of Certificate\*\*\*

Hemraj  
Calibrated By

Hemraj

Calibration Engineer

Reviewed & Approved By  
L. V. Mahato  
Chief Executive Officer (CEO)



Note: - 1. The calibration results reported in this certificate are valid at the time of and stated condition of measurement.

- This report is not to be reproducing wholly or in part and cannot be used as evidence in the Court of Law and should not be used in any advertising media without our special permission in writing.
- Total liability of our Organization is limited to the Invoiced amount.
- Calibration Certificate without signature are not valid.
- All our certificates are metrological traceable.
- The results relate only to the item calibrated.
- UUC: Unit Under Calibration.
- Expanded Uncertainty at 95 % confidence level with coverage factor K=2.
- Scan this QR code to download our NABL Accreditation Scope & Certificate.



Working : W.Z.1390/Z-2, 1st Floor, Nangal Raya Extn., Pankha Road, New Delhi-110046

Regd. Office : R.Z.56A, Indra Park, Gali No.25, Palam Colony, New Delhi-110045

M.: 9958445875, 9313768426, 9999761885 E-Mail : uniquecalibrationservices@gmail.com