

VIEW LAB MONTHLY SUMMARY

Lab Name	SADAR HOSPITAL		Details About Robust Analysis							
Lab No	18474		Detail About Monthly Summary							
Month	November		Detail about SDI							
Year	2023									
Constituent Group	Chemistry II									
Click on the analyte to view Graphical Data			All Analyser Result		Print					
			Date of Result Entered : 08/11/2023							
			Date of Report Published : 05/12/2023							
Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants CV	SD	Your Value	SDI	U
1	GLUCOSE II	GOD-POD II	Any Analyser (Automation / Semi Automation)	1594	238.88	9.59	22.91	211.2 mg/dL	-1.21	1.15
2	UREA II	Urease UV / GLDH II	Any Analyser (Automation / Semi Automation)	1098	67.14	11.47	7.70	84 mg/dL	2.19	0.46
3	CREATININE II	Jaffes Kinetic- Alkaline Picrate II	Any Analyser (Automation / Semi Automation)	1210	4.52	12.57	0.57	4.35 mg/dL	-0.30	0.03
4	URIC ACID II	Enzymatic / Uricase Colorimetric II	Any Analyser (Automation / Semi Automation)	1323	8.13	12.01	0.98	9.2 mg/dL	1.10	0.05
5	CHOLESTROL II	CHOD-PAP II	Any Analyser (Automation / Semi Automation)	1480	98.77	11.12	10.98	102 mg/dL	0.29	0.57
6	TRIGLYCERIDE II	GPO-PAP / Enzymatic Colorimetric / End Point II	Any Analyser (Automation / Semi Automation)	1379	88.73	18.66	16.56	88 mg/dL	-0.04	0.89
SDI Range			Interpretation							
Within -1.00 to +1.00			Excellent.							
Within ±1.01 to ±2.00			Good.							
Within ±2.01 to ±2.99			Accept with caution. Warning Signal.							
Beyond ±3.0			Unacceptable performance. Action Signal.							

Links

Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Sign-out