



# CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY  
CMC EXTERNAL QUALITY ASSURANCE SCHEME  
MONTHLY SUMMARY REPORT - OCTOBER 2023



Lab Name **DIVINE TOUCH MEDI CLINIC** Lab No **16585**  
Constituent Group **Chemistry II** Date of Result Entered : **19/10/2023**  
PT item **Lyophilized human serum based** Date of Report Published : **04/11/2023**

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE II	GOD-POD II	Agape	566	131.74	9.55	12.58	131.73 mg/dL	0.00	1.06
2	UREA II	Urease UV / GLDH II	Agape	476	86.59	9.95	8.61	92.65 mg/dL	0.70	0.79
3	CREATININE II	Jaffes Kinetic- Alkaline Picrate II	Agape	402	6.62	9.56	0.63	8.46 mg/dL	2.91	0.06
4	T.BILIRUBIN II	Diazonium Salt ( Colorimetric ) / Jendrassik II	Agape	449	8.58	10.71	0.92	8.39 mg/dL	-0.21	0.09
5	T-PROTEIN II	Biuret - Colorimetric II	Agape	314	5.47	7.00	0.38	5.39 g/dL	-0.21	0.04
6	ALBUMIN II	BCG - Colorimetric II	Agape	372	3.23	9.35	0.30	3.42 g/dL	0.63	0.03
7	URIC ACID II	Enzymatic / Uricase Colorimetric II	Agape	417	4.87	15.03	0.73	4.46 mg/dL	-0.56	0.07
8	CHOLESTROL II	CHOD-PAP II	Agape	529	107.32	11.49	12.33	81.93 mg/dL	-2.06	1.07
9	TRIGLYCERIDE II	GPO-PAP / Enzymatic Colorimetric / End Point II	Agape	381	96.29	15.53	14.95	85.72 mg/dL	-0.71	1.53

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within $\pm 1.01$ to $\pm 2.00$	Good.
Within $\pm 2.01$ to $\pm 2.99$	Accept with caution. Warning Signal.
Beyond $\pm 3.0$	Unacceptable performance. Action Signal.

LAB ADDRESS :  
DIVINE TOUCH MEDI CLINIC  
COLONEL CHOWMUHANI, AGARTALA  
TRIPURA  
TRIPURA799001

Coordinator Contact Details:  
Email:clinqc@cmcvellore.ac.in  
Contact Number: 0416-2283102

*Pamela Christudoss*  
Dr. Pamela Christudoss  
CMC EQAS Coordinator  
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.  
Data in CMC EQAS reports is confidential  
CMC EQAS does not sub contract any components  
\*\*\*\*\* End of Report \*\*\*\*\*