



# TMSCC

TESTING MACHINE SERVICE AND CALIBRATION CENTRE  
Precision is Our Destination.....

An ISO 9001 : 2015 Certified company  
Calibration of Various Instruments & Testing Machines...



CC - 3125

Form No.: TMSCCR/23						
CALIBRATION CERTIFICATE						
CALIBRATION CERTIFICATE NO.: 2023/07/261						Page 1 of 1
ISSUE DATE : 13-Jul-2023						
ULR No. : CC312523000009527P						
1.0	ISSUED TO :		M/s. South Calcutta Diagnostic Center 570, Tentulberia Road, P.O.: Garla, P.S.: Sonarpur, Kolkata - 700084			
1.1	Service Request Form No.:		SRF/2023/07/03/03			
1.2	Service Request Date :		03-07-2023			
1.3	Location :		AI Lab			
1.4	Description identification of item to be calibrated :					
	i	Name :	Digital Thermo Hygrometer	ii	Make :	HTC
	iii	Model / Type No.:	HTC-2	iv	S/L No.:	--
	v	LD.No.:	SCDC/DTHM/02	vi	Job Code No.:	2023/07/261
	vii	Range :	[(-)50 to 70] °C & (10 to 99)% RH	viii	Resolution :	0.1°C & 1% RH
	ix	Accuracy :	As Per DKD R5 - 1 & DKD-R5-7	x	End User :	--
1.5	Full / Partial Calibration :		Partial Calibration			
1.6	Applicable specification of item to be calibrated: Accuracy / permissible limit :					Not Specified.
1.7	Date of receipt of item :		3-Jul-2023	1.8	Date of calibration :	
					3-Jul-2023	
1.9	Calibration due on :		2-Jul-2024	2.0	Frequency of calibration once in :	
					12 Months	
2.1	Environmental condition during calibration :		Temperature : 25.1°C			
			Humidity : 53% RH			
2.2	Basis of calibration :		SOP/10.02, SOP/10.04			
2.3	Traceability : Standards used for calibration are traceable to National Standards through NABL accredited Laboratory.					
	Name of Instrument		SL No. / Id no.	Certificate No.	Lab Certificate No.	Calibrated on
	Temp. & Humidity Probe with Indicator		(SL No. : No.- 2022C01005/ 20571921)	ATL/T/090323/001	CC-2590	06-03-2023
	Digital Temperature Indicator with Sensor(RTD)		(SL No.- 18K588073) (Sensor SL No.- 19102403)	TSC/22-23/12095-30	CC - 2231	17-11-2022
Due on						
						06-03-2024
						17-11-2023
CALIBRATION RESULTS						
Sl. No.	Ref. Bath set in °C	Observed Reading at DUC in °C (Avg. of five readings)	Observed Reading at Ref. Std. in °C (Avg. of five readings)	Error in °C	Uncertainty in ± °C	Acceptance Criteria
1	-35.0	-35.0	-34.961	0.039	0.18	Pass
2	-5.0	-5.0	-4.994	0.006	0.18	Pass
3	10.0	10.0	9.989	-0.011	0.18	Pass
4	30.0	30.0	29.967	-0.033	0.18	Pass
5	70.0	70.0	69.923	-0.077	0.31	Pass
Maximum Permissible Error : ± 0.3% of rdg						
Sl. No.	Ref. Source set in %	Observed Reading at DUC in % (Avg. of five readings)	Observed Reading at Ref. Std. in % (Avg. of five readings)	Error in %	Measurement Uncertainty in ± %	Acceptance Criteria
1	25	25	25.1	0.1	1.3	Pass
2	50	50	50.3	0.3	1.3	Pass
3	90	90	91.2	1.2	1.3	Pass
Maximum Permissible Error : ± 3% of rdg						
Measurement Uncertainty at 95% Confidence Level where Coverage Factor k = 2						
REMARKS : The DUC has been calibrated over its range. The readings observed are tabulated above. The reference standard is traceable to National standard.						
DUC: Device Under Calibration.						
Physical Status of the DUC : OK						

Calibrated By :

(Calibration Engineer)

Checked By :

(Technical Manager)

Approved By :

(Sr. Calibration Engineer)

END OF CERTIFICATE

