

AMS Healthcare

(Erstwhile : Ace Medical Solutions (India) Pvt. Ltd.)

An ISO 9001-2015 Certified Company

No. 770/1, Jothi Complex 2nd Floor, Alagar Kovil Main Road, Surveyor Colony,
Madurai - 625 007, Tamilnadu, India Mobile : +91 94434 74610



GSTIN. 33AOSPA9153Q1ZA • DL No. TN-12-20B -00398, TN-12-21B-00398

CALIBRATION CERTIFICATE

Customer Name & Address:	M/S. SUN Clinical Laboratory, #424, Mangammal Salai, K.T.C. Nagar, Tirunelveli.	Certificate No:	AMS/HSP/289/5184
		Issuing Date:	30/10/2023
		Page No:	1 of 1

Details of Device under Calibration (DUC):

DUC:	Micro Pipette	Id. No:	SC/LAB/FMP/002
Make:	Astra	Department:	Laboratory
Model:	AMP 103LX	Calibrated On:	28/10/2023
Sl. No:	OX61978	Calibration Due Suggested:	27/10/2024
Type:	Fixed	DUC Condition on Receipt:	OK
Range & Least Count:	50µl		

Traceability Details:

Sl.No.	Test Equipment	Make	Certificate Number	Valid Up to
1	Electronic Weighing Scale	EWC/SF400C	PC/01/23/ATL/0512-02	31/01/2024
2	Thermo Hygrometer	Beetech/088	PC/01/23/ATL/0210-02	31/01/2024

Room Temperature: 25°C±2

Room Humidity: 30-60%

(At the time of Calibration)

Test Results:

SL. No.	DUC Reading (µl)	Standard Reading (µl)		Correction (µl)
		Trial 1	Trial 2	
1	50	50.00	50.00	0

Allowable Tolerance: User Defined

Note:

1. Calibration Points were selected as per customer request.
2. Each reading were taken by Repeatability Test (i.e.) each reading in a row is a mean of 10 readings.
3. Results Reported are valid at the time of and under the stated conditions of measurements.



Calibrated By:
(Calibration Engineer)

Approved by:
(Technical Manager)

aceagm@gmail.com

amshealthcareindia@gmail.com

94434 74610, 93455 56148

