

CERTIFICATE OF CALIBRATION

Customer Name & Address	Community Health Centre Moth Jhansi Uttar Pradesh 284303 India	Certificate No.	Z3JMTCHC20
		Date of Calibration	28.02.2023
		Valid Upto	27.02.2024
		Date of Issue	01.03.2023

Instruments Details			
Instrument Name	CENTRIFUGE	S No.	NA
Make/Model No.	Remi/R-4C	Location	LABORATORY
ID No.	MHUP - 402-0104	Visual Inspection	OK

Details of Reference Standards & Major Equipments Used			
Equipment Name	ELECTRICAL SAFETY ANALYZER	DIGITAL TACHO METER	THERMO HYGROMETER
Make	FLUKE	METRAVI	HTC
Model/SR No.	ESA612/2959012	NCTM 1000/190518478	288-ATH/NA
Calibration Validity	23/01/2024	20/07/2023	20/07/2023
Calibration By	STIC	UNIQUE CALIBRATION SOLUTIONS	UNIQUE CALIBRATION SOLUTIONS

Environmental Condition			
Temperature	27 °C	Relative Humidity	56 %

Calibration Results				
SL NO	SET READING (RPM)	MEASURED READING (RPM)	TOLERANCE(°C)	STATUS
1	1000	980	±3%	P
2	2000	2002	±3%	P
3	2500	2535	±3%	P

REMARKS:-

- ❖ This result of calibration refers to only to the particular items submitted for calibration.
- ❖ The above result is valid at the time of and under the stated conditions measurement.
- ❖ Calibration certificate issued for weight & measure parameters like mass, weighting balance, volumetric equipment, measuring scales/tapes etc. Are scientific purpose only and should not used for commercial.

Calibrated By

KHAN AMIR



Approved By

ASWANTH TM

