

CERTIFICATE OF CALIBRATION

Customer Name & Address	Community Health Centre Moth Jhansi Uttar Pradesh 284303 India	Certificate No.	Z3JMTCHC22
		Date of Calibration	28.02.2023
		Valid Upto	27.02.2024
		Date of Issue	01.03.2023

Instruments Details			
Instrument Name	Refrigerator	S No.	na
Make/Model No.	Whirlpool/icemagic	Location	ICTC DEPT.
ID No.	MHUP - 402-0570	Visual Inspection	OK

Details of Reference Standards & Major Equipments Used			
Equipment Name	THERMO HYGROMETER	ELECTRICAL SAFETY ANALYZER	DIGITAL THERMOMETER
Make	HTC	FLUKE	METRAVI
Model/SR No.	288-ATH/NA	ESA612/2959012	DTM-900/200923407
Calibration Validity	20/07/2023	23/01/2024	20/07/2023
Calibration By	UNIQUE CALIBRATION SOLUTIONS	STIC	UNIQUE CALIBRATION SOLUTIONS

Environmental Condition			
Temperature	27 °C	Relative Humidity	56 %

Calibration Results				
SL NO	SET READING (°C)	MEASURED READING(°C)	TOLERANCE(°C)	STATUS
1	4	4.1	±2	P
2	4	3.4	±2	P
3	4	3.9	±2	P

ELECTRICAL SAFETY REPORT ATTACHED

REMARKS:-

- ❖ This result of calibration refers to only to the particular items submitted for calibration.
- ❖ The above result is valid at the time of and under the stated conditions measurement.
- ❖ Calibration certificate issued for weight & measure parameters like mass, weighting balance, volumetric equipment, measuring scales/tapes etc. Are scientific purpose only and should not used for commercial.

Calibrated By

KHAN AMIR



Approved By

ASWANTH TM

