

Ace Medical Solutions (India) Pvt. Ltd.

An ISO 9001-2015 Certified Company

No. 770/1, Jothi Complex 2nd Floor, Alagar Kovil Main Road, Surveyor Colony,
Madurai - 625 007, Tamilnadu, India Mobile : +91 94434 74610



GSTIN. 33ADSPA9153Q1ZA • DL No. TN-12-20B -00398, TN-12-21B-00398

CALIBRATION CERTIFICATE

Customer Name & Address:	M/S. Care Diagnostics & Poly Clinic, #4/442-D2, N.G.G.O. Colony, Bagalur Road, Hosur-635109	Certificate No:	AMS/LAB/289/2607
		Issuing Date:	08/12/2022
		Page No:	1 of 2

Details of Device under Calibration (DUC):

DUC:	HBA1C Analyser	Id. No:	CDP/LAB/HBA1C/001
Make:	SD Biosensor	Department:	Laboratory
Model:	Standard F200	Calibrated On:	06/12/2022
Sl. No:	FA20EBITG039	Calibration Due Suggested:	05/12/2023
Type:	Semi Auto		
Range & Least Count:	---	DUC Condition on Receipt:	OK

Traceability Details:

Sl.No.	Test Equipment	Make	Certificate Number	Valid Up to
1	Electrical Safety Analyser	RigelSafeTest60	PSG/CIRD/BMCC/6908/2022	31/01/2023
2	Manometer	Lutron/PM9100	PC/01/22/ATL/0432-05	30/01/2023
3	Electronic Weighing Scale	EWC/SF400C	PC/01/22/ATL/0432-02	30/01/2023
4	Thermo Hygrometer	Beetech/088	PC/01/22/ATL/0265-01	30/01/2023

Room Temperature: 25°C±2 Room Humidity: 30-60% (At the time of Calibration)

ESA Test Results:

SL.No	Parameter	ESA Reading	Unit	Normal Range	Condition
1	Current	200	mA	500	Normal
2	Frequency	50	Hz	49 -50	Normal
3	Mains Voltage Line - Neutral	219.7	Volts	220 - 240	Normal
4	Mains Voltage Line - Earth	113.7	Volts	220 - 240	Normal
5	Mains Voltage Neutral - Earth	106.4	Volts	< 7	Normal

*Electrical Safety: Please check the attached log sheet

SL.No	Parameter	Measured	Standard	Correction	Unit
1	Lamp Voltage	11.78	12.0	-0.22	Volts
2	Detector Sensitivity	0.0	0.0	0.0	% Absorbance
3	Sipper Volume	432.5	450	-17.5	ml
4	Peristaltic Pump pressure	-14.42	-15.0	-0.58	cmH2O

Allowable Tolerance: set by Customer

Note:

1. Calibration Points were selected as per customer request.
2. Each reading was taken by Repeatability Test (i.e.) each reading in a row is a mean of 10 readings.
3. Results Reported are valid at the time of and under the stated conditions of measurements.

Calibrated By:
(Calibration Engineer)

aceagm@gmail.com



amshealthcareindia@gmail.com

Approved by:
(Technical Manager)

94434 74610, 93455 56148

Ace Medical Solutions (India) Pvt. Ltd.

An ISO 9001-2015 Certified Company

No. 770/1, Jothi Complex 2nd Floor, Alagar Kovil Main Road, Surveyor Colony,
Madurai - 625 007, Tamilnadu, India Mobile : +91 94434 74610



GSTIN. 33ADSPA9153Q17A • DL No. TN-12-20B-00398, TN-12-21B-00398

CALIBRATION CERTIFICATE

Customer Name & Address:	M/S. Care Diagnostics & Poly Clinic, #4/442-D2, N.G.G.O. Colony, Bagalur Road, Hosur-635109	Certificate No:	AMS/LAB/289/2605
		Issuing Date:	08/12/2022
		Page No:	1 of 2

Details of Device under Calibration (DUC):

DUC:	Electrolyte Analyser	Id. No:	CDP/LAB/EA/001
Make:	Arkray	Department:	Laboratory
Model:	Spot Chem-L SE-1520	Calibrated On:	06/12/2022
Sl. No:	122208017	Calibration Due Suggested:	05/12/2023
Type:	Semi Auto	DUC Condition on Receipt:	OK
Range & Least Count:	---		

Traceability Details:

Sl.No.	Test Equipment	Make	Certificate Number	Valid Up to
1	Electrical Safety Analyser	RigelSafeTest60	PSG/CIRD/BMCC/6908/2022	31/01/2023
2	Manometer	Lutron/PM9100	PC/01/22/ATL/0432-05	30/01/2023
3	Electronic Weighing Scale	EWC/SF400C	PC/01/22/ATL/0432-02	30/01/2023
4	Thermo Hygrometer	Beetech/088	PC/01/22/ATL/0265-01	30/01/2023

Room Temperature: 25°C±2

Room Humidity: 30-60%

(At the time of Calibration)

ESA Test Results:

SL.No	Parameter	ESA Reading	Unit	Normal Range	Condition
1	Current	200	mA	500	Normal
2	Frequency	50	Hz	49 -50	Normal
3	Mains Voltage Line - Neutral	219.7	Volts	220 - 240	Normal
4	Mains Voltage Line - Earth	113.7	Volts	220 - 240	Normal
5	Mains Voltage Neutral - Earth	106.4	Volts	< 7	Normal

*Electrical Safety: Please check the attached log sheet

SL.No	Parameter	Measured	Standard	Correction	Unit
1	Lamp Voltage	11.78	12.0	-0.22	Volts
2	Detector Sensitivity	0.0	0.0	0.0	% Absorbance
3	Sipper Volume	432.5	450	-17.5	ml
4	Peristaltic Pump pressure	-14.42	-15.0	-0.58	cmH ₂ O

Allowable Tolerance: set by Customer



Note:

1. Calibration Points were selected as per customer request.
2. Each reading was taken by Repeatability Test (i.e.) each reading in a row is a mean of 10 readings.
3. Results Reported are valid at the time of and under the stated conditions of measurements.

Calibrated By:
(Calibration Engineer)



Approved by:
(Technical Manager)

Ace Medical Solutions (India) Pvt. Ltd.

An ISO 9001-2015 Certified Company

No. 770/1, Jothi Complex 2nd Floor, Alagar Kovil Main Road, Surveyor Colony,
Madurai - 625 007, Tamilnadu, India Mobile : +91 94434 74610



GSTIN. 33ADSPA9153Q1ZA • DL No. TN-12-20B -00398, TN-12-21B-00398

CALIBRATION CERTIFICATE

Customer Name & Address:	M/S. Care Diagnostics & Poly Clinic, #4/442-D2, N.G.G.O. Colony, Bagalur Road, Hosur-635109	Certificate No:	AMS/LAB/289/2609
		Issuing Date:	08/12/2022
		Page No:	1 of 2

Details of Device under Calibration (DUC):

DUC:	Fully Automated Analyser	Id. No:	CDP/LAB/FAA/001
Make:	Arkray	Department:	Laboratory
Model:	CKK100	Calibrated On:	06/12/2022
Sl. No:	24003	Calibration Due Suggested:	05/12/2023
Type:	Fully Auto	DUC Condition on Receipt:	OK
Range & Least Count:	---		

Traceability Details:

Sl.No.	Test Equipment	Make	Certificate Number	Valid Up to
1	Electrical Safety Analyser	RigelSafeTest60	PSG/CIRD/BMCC/6908/2022	31/01/2023
2	Manometer	Lutron/PM9100	PC/01/22/ATL/0432-05	30/01/2023
3	Electronic Weighing Scale	EWC/SF400C	PC/01/22/ATL/0432-02	30/01/2023
4	Thermo Hygrometer	Beetech/088	PC/01/22/ATL/0265-01	30/01/2023

Room Temperature: 25°C±2

Room Humidity: 30-60%

(At the time of Calibration)

ESA Test Results:

SL.No	Parameter	ESA Reading	Unit	Normal Range	Condition
1	Current	200	mA	500	Normal
2	Frequency	50	Hz	49 -50	Normal
3	Mains Voltage Line - Neutral	219.7	Volts	220 - 240	Normal
4	Mains Voltage Line - Earth	113.7	Volts	220 - 240	Normal
5	Mains Voltage Neutral - Earth	106.4	Volts	< 7	Normal

*Electrical Safety: Please check the attached log sheet

SL.No	Parameter	Measured	Standard	Correction	Unit
1	Lamp Voltage	11.78	12.0	-0.22	Volts
2	Detector Sensitivity	0.0	0.0	0.0	% Absorbance
3	Sipper Volume	432.5	450	-17.5	ml
4	Peristaltic Pump pressure	-14.42	-15.0	-0.58	cmH ₂ O

Allowable Tolerance: set by Customer

Note:

1. Calibration Points were selected as per customer request.
2. Each reading was taken by Repeatability Test (i.e.) each reading in a row is a mean of 10 readings.
3. Results Reported are valid at the time of and under the stated conditions of measurements.

Calibrated By:
(Calibration Engineer)



Approved by:
(Technical Manager)

Ace Medical Solutions (India) Pvt. Ltd.

An ISO 9001-2015 Certified Company

No. 770/1, Jothi Complex 2nd Floor, Alagar Kovil Main Road, Surveyor Colony,
Madurai - 625 007, Tamilnadu, India Mobile : +91 94434 74610



GSTIN. 33ADSPA9153Q1ZA • DL No. TN-12-20B -00398, TN-12-21B-00398

CALIBRATION CERTIFICATE

Customer Name & Address:	M/S. Care Diagnostics & Poly Clinic, #4/442-D2, N.G.G.O. Colony, Bagalur Road, Hosur-635109	Certificate No:	AMS/LAB/289/2602
		Issuing Date:	08/12/2022
		Page No:	1 of 1

Details of Device under Calibration (DUC):

DUC:	Centrifuge	Id. No:	CDP/LAB/CEN/01
Make:	Remi	Department:	Laboratory
Model:	R-4c	Calibrated On:	06/12/2022
Sl. No:	CBLC-3023	Calibration Due Suggested:	05/12/2023
Type:	Manual	DUC Condition on Receipt:	OK
Range & Least Count:	0 - 3500RPM		

Traceability Details:

Sl.No.	Test Equipment	Make	Certificate Number	Valid Up to
1	Digital Tachometer	Mextech / DT-2234C	PC/01/22/ATL/0432-01	30/01/2023
2	Thermo Hygrometer	Beetech/088	PC/01/22/ATL/0265-01	30/01/2023
3	Stop Watch	Nivia / JS 307	PC/01/22/ATL/0134-01	30/01/2023

Room Temperature: 25°C±2

Room Humidity: 30-60%

(At the time of Calibration)

Test Results:

SL.No.	DUC Reading (Positions) RPM	Standard Reading (RPM)	Correction (RPM)
1	1000	1020	+260
2	2000	2040	-121
3	2500	2580	-120
4	3500	3010	-110
5	3500	3410	-390

Allowable Tolerance: Set by Customer



Note:

1. Calibration Points were selected as per customer request.
2. Each reading was taken by Repeatability Test (i.e.) each reading in a row is a mean of 10 readings.
3. Results Reported are valid at the time of and under the stated conditions of measurements.

Calibrated By:
(Calibration Engineer)



Approved by:
(Technical Manager)

aceagm@gmail.com

amshealthcareindia@gmail.com

94434 74610, 93455 56148

Ace Medical Solutions (India) Pvt. Ltd.

An ISO 9001-2015 Certified Company

No. 770/1, Jothi Complex 2nd Floor, Alagar Kovil Main Road, Surveyor Colony,
Madurai - 625 007, Tamilnadu, India Mobile : +91 94434 74610



GSTIN. 33ADSPA9153Q1ZA • DL No. TN-12-20B-00398, TN-12-21B-00398

CALIBRATION CERTIFICATE

Customer Name & Address:	M/S. Care Diagnostics & Poly Clinic, #4/442-D2, N.G.G.O. Colony, Bagalur Road, Hosur-635109	Certificate No:	AMS/LAB/289/2610
		Issuing Date:	08/12/2022
		Page No:	1 of 2

Details of Device under Calibration (DUC):

DUC:	Haematology Analyser-Cell Counter	Id. No:	CDP/LAB/HA/002
Make:	Mindray	Department:	Laboratory
Model:	BC-5130	Calibrated On:	06/12/2022
Sl. No:	TR-1C006435	Calibration Due Suggested:	05/12/2023
Type:	Semi Auto	DUC Condition on Receipt:	OK
Range & Least Count:	---		

Traceability Details:

Sl.No.	Test Equipment	Make	Certificate Number	Valid Up to
1	Electrical Safety Analyser	RigelSafeTest60	PSG/CIRD/BMCC/6908/2022	31/01/2023
2	Manometer	Lutron/PM9100	PC/01/22/ATL/0432-05	30/01/2023
3	Electronic Weighing Scale	EWC/SF400C	PC/01/22/ATL/0432-02	30/01/2023
4	Thermo Hygrometer	Beetech/088	PC/01/22/ATL/0265-01	30/01/2023

Room Temperature: 25°C±2

Room Humidity: 30-60%

(At the time of Calibration)

ESA Test Results:

SL.No	Parameter	ESA Reading	Unit	Normal Range	Condition
1	Current	300	mA	500	Normal
2	Frequency	49.7	Hz	49 -50	Normal
3	Mains Voltage Line - Neutral	230.7	Volts	220 - 240	Normal
4	Mains Voltage Line - Earth	234.5	Volts	220 - 240	Normal
5	Mains Voltage Neutral - Earth	6.5	Volts	< 7	Normal

*Electrical Safety: Please check the attached log sheet

SL.No	Parameter	Measured	Standard	Correction	Unit
1	Lamp Voltage	11.78	12.0	-0.22	Volts
2	Detector Sensitivity	0.0	0.0	0.0	% Absorbance
3	Sipper Volume	432.5	450	-17.5	
4	Peristaltic Pump pressure	-14.42	-15.0	-0.58	cmH ₂ O

Allowable Tolerance: set by Customer

Note:

1. Calibration Points were selected as per customer request.
2. Each reading was taken by Repeatability Test (i.e.) each reading in a row is a mean of 10 readings.
3. Results Reported are valid at the time of and under the stated conditions of measurements.

Calibrated By:
(Calibration Engineer)



Approved by:
(Technical Manager)

Ace Medical Solutions (India) Pvt. Ltd.

An ISO 9001-2015 Certified Company

No. 770/1, Jothi Complex 2nd Floor, Alagar Kovil Main Road, Surveyor Colony,
Madurai - 625 007, Tamilnadu, India Mobile : +91 94434 74610



GSTIN: 33ADSPA9153Q1ZA • DL No. TN-12-20B -00398, TN-12-21B-00398

CALIBRATION CERTIFICATE

Customer Name & Address:	M/S. Care Diagnostics & Poly Clinic, #4/442-D2, N.G.G.O. Colony, Bagalur Road, Hosur-635109	Certificate No:	AMS/LAB/289/2608
		Issuing Date:	08/12/2022
		Page No:	1 of 2

Details of Device under Calibration (DUC):

DUC:	Serum Protein Analyser	Id. No:	CDP/LAB/SPA/001
Make:	Tulip Diagnostics	Department:	Laboratory
Model:	Turbodyne- SC	Calibrated On:	06/12/2022
Sl. No:	P15070788	Calibration Due Suggested:	05/12/2023
Type:	Semi Auto	DUC Condition on Receipt:	OK
Range & Least Count:	---		

Traceability Details:

Sl.No.	Test Equipment	Make	Certificate Number	Valid Up to
1	Electrical Safety Analyser	RigelSafeTest60	PSG/CIRD/BMCC/6908/2022	31/01/2023
2	Manometer	Lutron/PM9100	PC/01/22/ATL/0432-05	30/01/2023
3	Electronic Weighing Scale	EWC/SF400C	PC/01/22/ATL/0432-02	30/01/2023
4	Thermo Hygrometer	Beetech/088	PC/01/22/ATL/0265-01	30/01/2023

Room Temperature: 25°C±2 Room Humidity: 30-60% (At the time of Calibration)

ESA Test Results:

SL.No	Parameter	ESA Reading	Unit	Normal Range	Condition
1	Current	200	mA	500	Normal
2	Frequency	50	Hz	49 -50	Normal
3	Mains Voltage Line - Neutral	219.7	Volts	220 - 240	Normal
4	Mains Voltage Line - Earth	113.7	Volts	220 - 240	Normal
5	Mains Voltage Neutral - Earth	106.4	Volts	< 7	Normal

*Electrical Safety: Please check the attached log sheet

SL.No	Parameter	Measured	Standard	Correction	Unit
1	Lamp Voltage	11.78	12.0	-0.22	Volts
2	Detector Sensitivity	0.0	0.0	0.0	% Absorbance
3	Sipper Volume	432.5	450	-17.5	ml
4	Peristaltic Pump pressure	-14.42	-15.0	-0.58	cmH ₂ O

Allowable Tolerance: set by Customer

Note:

1. Calibration Points were selected as per customer request.
2. Each reading was taken by Repeatability Test (i.e.) each reading in a row is a mean of 10 readings.
3. Results Reported are valid at the time of and under the stated conditions of measurements.

Calibrated By:
(Calibration Engineer)



Approved by:
(Technical Manager)

Ace Medical Solutions (India) Pvt. Ltd.

An ISO 9001-2015 Certified Company

No. 770/1, Jothi Complex 2nd Floor, Alagar Kovil Main Road, Surveyor Colony,
Madurai - 625 007, Tamilnadu, India Mobile : +91 94434 74610



GSTIN: 33ADSPA9153Q1ZA • DL No. TN-12 20B 00398, TN-12 21B-00398

CALIBRATION CERTIFICATE

Customer Name & Address:	M/S. Care Diagnostics & Poly Clinic, #4/442-D2, N.G.G.O. Colony, Bagalur Road, Hosur-635109	Certificate No:	AMS/LAB/289/2604
		Issuing Date:	08/12/2022
		Page No:	1 of 2

Details of Device under Calibration (DUC):

DUC:	Cell Counter	Id. No:	CDP/LAB/CC/001
Make:	Horiba	Department:	Laboratory
Model:	Micros-60	Calibrated On:	06/12/2022
Sl. No:	6050T95992	Calibration Due Suggested:	05/12/2023
Type:	Semi Auto	DUC Condition on Receipt:	OK
Range & Least Count:	---		

Traceability Details:

Sl.No.	Test Equipment	Make	Certificate Number	Valid Up to
1	Electrical Safety Analyser	RigelSafeTest60	PSG/CIRD/BMCC/6908/2022	31/01/2023
2	Manometer	Lutron/PM9100	PC/01/22/ATL/0432-05	30/01/2023
3	Electronic Weighing Scale	EWC/SF400C	PC/01/22/ATL/0432-02	30/01/2023
4	Thermo Hygrometer	Beetech/088	PC/01/22/ATL/0265-01	30/01/2023

Room Temperature: 25°C±2 Room Humidity: 30-60%

(At the time of Calibration)

ESA Test Results:

SL.No	Parameter	ESA Reading	Unit	Normal Range	Condition
1	Current	200	mA	500	Normal
2	Frequency	49.8	Hz	49 - 50	Normal
3	Mains Voltage Line - Neutral	230.7	Volts	220 - 240	Normal
4	Mains Voltage Line - Earth	235.3	Volts	220 - 240	Normal
5	Mains Voltage Neutral - Earth	8.2	Volts	< 7	Normal

*Electrical Safety: Please check the attached log sheet

SL.No	Parameter	Measured	Standard	Correction	Unit
1	Lamp Voltage	11.78	12.0	-0.22	Volts
2	Detector Sensitivity	0.0	0.0	0.0	% Absorbance
3	Sipper Volume	432.5	450	-17.5	ml
4	Peristaltic Pump pressure	-14.42	-15.0	-0.58	cmH ₂ O

Allowable Tolerance: set by Customer

Note:

1. Calibration Points were selected as per customer request.
2. Each reading was taken by Repeatability Test (i.e.) each reading in a row is a mean of 10 readings.
3. Results Reported are valid at the time of and under the stated conditions of measurements.

Calibrated By:
(Calibration Engineer)

aceagm@gmail.com



amshealthcareindia@gmail.com

Approved by:
(Technical Manager)

94434 74610, 93455 56148

Ace Medical Solutions (India) Pvt. Ltd.

An ISO 9001-2015 Certified Company

No. 770/1, Jothi Complex 2nd Floor, Alagar Kovil Main Road, Surveyor Colony,
Madurai - 625 007, Tamilnadu, India Mobile : +91 94434 74610



GSTIN. 33ADSPA9153Q1ZA • DL No. TN-12-20B -00398, TN-12-21B-00398

CALIBRATION CERTIFICATE

Customer Name & Address:	M/S. Care Diagnostics & Poly Clinic, #4/442-D2, N.G.G.O. Colony, Bagalur Road, Hosur-635109	Certificate No:	AMS/LAB/289/2606
		Issuing Date:	08/12/2022
		Page No:	1 of 2

Details of Device under Calibration (DUC):

DUC:	Haemostasis Analyser- Coagulometer	Id. No:	CDP/LAB/HA/001
Make:	Hemostar	Department:	Laboratory
Model:	X.F-1.0	Calibrated On:	06/12/2022
Sl. No:	H1022S40B	Calibration Due Suggested:	05/12/2023
Type:	Semi Auto	DUC Condition on Receipt:	OK
Range & Least Count:	---		

Traceability Details:

Sl.No.	Test Equipment	Make	Certificate Number	Valid Up to
1	Electrical Safety Analyser	RigelSafeTest60	PSG/CIRD/BMCC/6908/2022	31/01/2023
2	Manometer	Lutron/PM9100	PC/01/22/ATL/0432-05	30/01/2023
3	Electronic Weighing Scale	EWC/SF400C	PC/01/22/ATL/0432-02	30/01/2023
4	Thermo Hygrometer	Beetech/088	PC/01/22/ATL/0265-01	30/01/2023

Room Temperature: 25°C±2 Room Humidity: 30-60% (At the time of Calibration)

ESA Test Results:

SL.No	Parameter	ESA Reading	Unit	Normal Range	Condition
1	Current	100	mA	500	Normal
2	Frequency	50	Hz	49 -50	Normal
3	Mains Voltage Line - Neutral	220.6	Volts	220 - 240	Normal
4	Mains Voltage Line - Earth	110.3	Volts	220 - 240	Normal
5	Mains Voltage Neutral - Earth	113.2	Volts	< 7	Normal

*Electrical Safety: Please check the attached log sheet

SL.No	Parameter	Measured	Standard	Correction	Unit
1	Lamp Voltage	11.78	12.0	-0.22	Volts
2	Detector Sensitivity	0.0	0.0	0.0	% Absorbance
3	Sipper Volume	432.5	450	-17.5	ml
4	Peristaltic Pump pressure	-14.42	-15.0	-0.58	cmH ₂ O

Allowable Tolerance: set by Customer

Note:

1. Calibration Points were selected as per customer request.
2. Each reading was taken by Repeatability Test (i.e.) each reading in a row is a mean of 10 readings.
3. Results Reported are valid at the time of and under the stated conditions of measurements.

Calibrated By:
(Calibration Engineer)



Approved by:
(Technical Manager)