



# Metropolis Healthcare Ltd., EQAS

POLO LABS PVT.LTD (617)

## Outlier And Analyte Summary Report

Outlier Details For Cycle No 230206 and Sample No 02

Report Date : 29/07/2023

Analyte	Instrument	Result Value	Standard Unit	Z-Score
Legend @ : z score $\leq$ 2.0 - Acceptable ! : 2.0 < z score < 3.0 - Warning X : z score $\geq$ 3 - Unacceptable # : Not Evaluated ⌚ : Delayed Result Entry * : Not considered for evaluation.				

**Problem Classification:** \_\_\_\_\_

\_\_\_\_\_  
\_\_\_\_\_

**Corrective Action:** \_\_\_\_\_

\_\_\_\_\_  
\_\_\_\_\_

**Reviewed by:** \_\_\_\_\_

**Dated:** \_\_\_\_\_

Instrument : Diasys Sys 200

Analyte	Standard Unit	Result Value	Mean	Z-Score	RMZ
@ ALT	U/L	7.00	6.00	1.3	0.80
@ Glucose	mg/dl	57.20	60.50	-0.81	0.47
@ Urea	mg/dl	17.30	17.70	0.39	0.58
@ Creatinine	mg/dl	0.70	0.80	0.41	0.32
@ Bilirubin Direct	mg/dl	0.08	0.09	-0.44	-0.73
@ Total Protein	g/dL	5.90	5.70	1.06	-0.08
@ Albumin	g/dL	3.30	3.53	0.36	1.46
@ Uric Acid	mg/dl	3.45	3.45	-0.89	-1.22
@ Calcium	mg/dl	6.90	7.17	0.19	0.08
@ Phosphorous Inorganic	mg/dl	3.40	3.40	0.00	0.58
@ Cholesterol	mg/dl	127.10	126.00	0.50	0.11
@ Triglycerides	mg/dl	106.40	106.47	-0.05	-0.25
@ HDL	mg/dl	30.30	33.50	0.59	-0.26
@ AST	U/L	11.00	11.33	0.15	0.06
@ ALP	U/L	42.00	39.00	-0.32	1.50
@ GGT	U/L	19.00	22.00	-1.66	-0.11
@ Amylase	U/L	46.00	48.87	-0.33	-0.02

Instrument : Siemens RAPIDLab 1240/1245/1260/1265

Analyte	Standard Unit	Result Value	Mean	Z-Score	RMZ
@ Chloride	mmol/L	166.00	--	1.55	0.09
@ Sodium	mmol/L	119.60	--	-0.78	1.26
@ Potassium	mmol/L	3.30	--	0.01	1.05

Legend @ : Acceptable

! : 2.0 < z score < 3.0 - Warning


X : z score ≥ 3 - Unacceptable

# : Not Evaluated

⌚ : Delayed Result Entry

\* : Not considered for evaluation.

Total Parameters	20
Not Evaluated Parameters	0
Evaluated Parameters	20
Outlier Parameters (X)	1
EQAS Score Biochemistry	95.00 %



**Dr Puneet Kumar Nigam**  
**PT coordinator & Technical Manager, MHL EQAS**

Unit No. 409-416.  
Commercial Building - 1A  
Kohinoor Mall, Kirol Road, Kurla (W),  
Mumbai - 400070