



Metropolis Healthcare Ltd., EQAS

POLO LABS PVT.LTD (617)

Outlier And Analyte Summary Report

Outlier Details For Cycle No 230206 and Sample No 05

Report Date : 29/07/2023

Analyte	Instrument	Result Value	Standard Unit	Z-Score
No Outlier				
Legend @ : z score \leq 2.0 - Acceptable ! : 2.0 < z score < 3.0 - Warning X : z score \geq 3 - Unacceptable # : Not Evaluated ⌚ : Delayed Result Entry * : Not considered for evaluation.				

Problem Classification: _____

Corrective Action: _____

Reviewed by: _____

Dated: _____

Instrument : Beckman Coulter DxH 500

Analyte	Standard Unit	Result Value	Mean	Z-Score	RMZ
@ Erythrocyte (RBC) Count	mill/cu.mm	2.34	2.32	-0.12	0.05
@ Total Leucocytes (WBC) count	/c.mm	3960.00	3966.57	-0.17	-0.16
@ Haemoglobin	g/dL	7.00	7.01	-0.23	-0.60
@ MCV (Mean Corpuscular Volume)	fL	92.20	92.04	0.02	0.00
@ MCH (Mean Corpuscular Hb)	pg	30.00	30.27	-0.17	-0.61
@ MCHC (Mean Corpuscular Hb Concentration)	g/dL	32.50	32.85	-0.14	-0.42
@ Platelet Count	1000/ μ L	143.10	148.14	-0.19	-0.06
@ Hematocrit	%	21.60	21.47	-0.07	0.07
@ RDW CV%	%	13.20	13.07	0.43	-0.01

Legend @ : Acceptable

! : 2.0 < z score < 3.0 - Warning


X : z score \geq 3 - Unacceptable

: Not Evaluated

⌚ : Delayed Result Entry

* : Not considered for evaluation.

Total Parameters	9
Not Evaluated Parameters	0
Evaluated Parameters	9
Outlier Parameters (X)	0
EQAS Score Haematology - CBC	100.00 %



Dr Puneet Kumar Nigam
PT coordinator & Technical Manager, MHL EQAS

Unit No. 409-416.

Commercial Building - 1A

Kohinoor Mall, Kirol Road, Kurla (W),

Mumbai - 400070