



# Metropolis Healthcare Ltd., EQAS

POLO LABS PVT.LTD (617)

## Outlier And Analyte Summary Report

Outlier Details For Cycle No 230206 and Sample No 01

Report Date : 29/07/2023

Analyte	Instrument	Result Value	Standard Unit
Legend @ : Acceptable X : Unacceptable # : Not Evaluated ⌚ Delayed Result Entry			

**Problem Classification:** \_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_

**Corrective Action:** \_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_

**Reviewed by:** \_\_\_\_\_

**Dated:** \_\_\_\_\_

Instrument : Manual

Analyte	Standard Unit	Result Value	Accepted Value
@ HIV Ag + Ab	S/CO	Reactive	Reactive
@ HBsAg	IU/mL	Non-reactive	Non-reactive
@ HCV	IU/mL	Non-reactive	Non-reactive
@ Rheumatoid Factor Qualitative	IU/mL	Normal	Normal
@ RPR (Syphilis Serology)	-	Non-reactive	Non-reactive
@ Typhidot IgM	-	Negative	Negative
@ ASO Qualitative	IU/mL	Normal	Normal

Instrument : J Mitra

Analyte	Standard Unit	Result Value	Accepted Value
# HIV Ab	S/CO	Non-reactive	

Legend @ : Acceptable  
 @ : Acceptable  
 X : Unacceptable  
 # : Not Evaluated  
 ⌚ : Delayed Result Entry

Total Parameters	8
Not Evaluated Parameters	1
Evaluated Parameters	7
Outlier Parameters (X)	0
EQAS Score Immunoassay Extended	92.0 %



**Dr Puneet Kumar Nigam**  
 PT coordinator & Technical Manager, MHL EQAS  
 Unit No. 409-416.  
 Commercial Building - 1A  
 Kohinoor Mall, Kirol Road, Kurla (W),  
 Mumbai - 400070