



Indian Calibration Services

An Exclusive House for Validation & Calibration of Analytical Instruments
 304 & 313, Laxmi Deep, District Centre, Near Nirman Vihar Metro Station, Laxmi Nagar, Delhi-110092
 Phone / Fax : (011) 22453259, 22434362, 9599482307



E-mail : indiancalibrationservices@gmail.com Website : www.multitechics.com

CALIBRATION CERTIFICATE

Page 1 of 2

CALIBRATION CERTIFICATE OF REFRIGERATOR

| | | | |
|------------------------------------|-------------------------|-------------------------------|------------|
| Calibration Certificate No. | 23100017583 | Issued On: | 29/08/2023 |
| ULR No. | CC214623100017583F | Calibration Date: | 26/08/2024 |
| Job Identification No. | ICS/C/T/08/17583 | Next Calibration Date: | 25/08/2024 |
| Ref. No. | SRF, Dated - 23/08/2023 | | |

| | |
|------------------------|---|
| CALIBRATED FOR: | M/s Polo Labs Pvt Ltd., C/o IVY Hospital, Khanna (Punjab) |
|------------------------|---|

EQUIPMENT DESCRIPTION

| | | | |
|-------------------|--------------|--------------------------|---------------|
| Name | Refrigerator | Visual Inspection | Ok |
| Make/Model | LG | I.D. No. | POLO/KH/RF/01 |
| Serial No. | | Least Count | On-Site |
| Range | 2 to 8°C | Calibration Site | On-Site |
| Location | Lab | | |

ENVIRONMENTAL CONDITIONS

| | | | |
|--------------------|------------------|-----------------|--------------|
| Temperature | 25.0 °C ± 5.0 °C | Humidity | 50 ± 15 % RH |
|--------------------|------------------|-----------------|--------------|

STANDARD REFERENCE DETAILS


| | | | |
|---|--------------------|------------------------|----------------------------|
| Traceable to National / International Standard | | | |
| Name of the Standard | Make | Certificate No. | Date of Calibration |
| Dig. Multichannel Data Logger | Microcubus | 23100000030 | 05/01/2023 |
| Agency/Traceability | Valid up to | | |
| Indian Calibration Services, Delhi | 05/01/2024 | | |

| | |
|--|---|
| USED EQUIPMENT DETAILS | N/A |
| PRINCIPLE/METHODOLOGY OF CALIBRATION: | As per Calibration Procedure No.: ICS-/CAL/SOP-TH04, ITS-90 (Comparison Method) |

RESULTS: Thermal Calibration (Temperature)
 All readings are in "°C"

| Set Value in °C | Measured Average Value in (°C) (Avg.) | | | | | | | | | Uncertainty(with 95% C.L. at k=2) (°C) |
|-------------------------|---------------------------------------|-----|-----|-----|-----|-----|-----|-----|-----|--|
| | S-1 | S-2 | S-3 | S-4 | S-5 | S-6 | S-7 | S-8 | S-9 | |
| 2 to 8 | 4.3 | 4.4 | 4.6 | 4.7 | 4.9 | 4.5 | 4.7 | 4.9 | 5.0 | ±1.1 |
| Max Stability:- | 0.4 | | | | | | | | | |
| Max Uniformity:- | 0.7 | | | | | | | | | |

Refrigerator



Remarks: - Calibration results are enclosed.

For and on behalf of
 Indian Calibration Services

- Note :**
- This report is not to be reproduced wholly or in part and cannot be used as an evidence in the Court of Law and should not be used in any advertising media without our special permission in writing.
 - The result listed refer only to the tested samples and applicable parameters. Endorsement of products is neither inferred nor implied.
 - Total liability of our Organisation is limited to the invoiced amount.
 - Samples will be destroyed after one month from the date of issue of Calibration Certificate unless otherwise specified.
 - In case any reconfirmation of contents of this Calibration Certificate is required. Please contact our office.
 - The calibration certificate/Test Report is valid only for the condition of the UUC at the time under stated condition of calibration.

Checked by **RAJESH KAPOOR**
 Authorised Signatory