

Ref: SNS/GK/23082 Date: 06/12/2023 Tel No: +91 9903209869 E-Mail Id:sns4.kolkata@gmail.com

Calibration & Validation Certificate: 23082

Customer Name : Meditrust Diagnostic Centre

33, Anukul Chandra Naskar Road

Garia

Kolkata - 700084

Instrument Model : Innolab 200

Serial No. : ESI032305PT121P

Date of Calibration : 06/12/2023

Validity : 05/06/2024

This is to certify that the above mentioned product has been verified of calibration electronically and the instrument is working satisfactory.

Next Calibration : **06/06/2024**


For SCIENCE & SERVICE

Gurudas Kusari

Engineer




Company's GSTIN: 19ACMFS3034L1ZO, PAN No.: ACMFS3034L



IEi'CEi'DE'8C7 | A9BHG :CF":BBC@56'&\$\$

ÖOCKŰÓÚŠ®ŰUŰÚŰ®}}[|æàG€€ Ü^|^æ•^kÖæc^kK€JÉ€HÉG€GG Ü^çi•i[}K€€ Ü^çi•i[}kÖæc^KÞ0Œ


Installation Qualification (IQ)	InnoLab ₂₀₀			
Company (Name & Add) Q-Line Biolech Pulltd New Delly-110028 Contact Person: My. Dipankar Chush Telephone No.: 7339521840	Customer (Name & Add) Meditury & Siagonstie Centre 33, Anukul Chandra Road, Garia-89, Contact Person Rahul Dll Telephone No. 9830411196			
Name of instrument: Innolab 200 (Semi Automatic Biochemistry Analyzer)				
Serial No. : ES1032305PT121P	Installation Date: 0.6 12 2023			
Perform the next steps Place the analyser on a Dry and Vibration free surface Check Power Supply AC 100-240 V (±10V), 50/60H; Check ground voltage	e motors and electrical contacts that are frequently turned on t. g inlet & outlet tubes if required ct to the power socket of instrument in back side			
	New Delhi John Pilling Charles Andrew Delhi			



IEi'CEi'DE'8C7 I A9BHG : CF"=BBC@56'&\$\$

ÖOCKŰÓÚŠMŰUŰÚŰM}}[|æåG€€ Ü^|^æ•^kÖæc^kK€JÉ€HÉG€GG Ü^çi•i[]K€€ Ü^çi•i[]kÖæc^KÞWE

Operational Qualification (OQ)




- Switch on the instrument on using power switch and wait 10 seconds, you will see the product name & Software version followed by main screen with eight different function dialogue boxes on display
- After self check, system displays main menu



- Check the functioning of touch screen by touching/operating it with hand and stylus provided.
- Test icon has sample, QC test, reagent blank and standard test functions
- View Result View test results after analysis, or history samples information. Query QC results, Standard results.

 Clear and backup the data.
- Reagent Setup Setup the assay parameters for analysis.
- QC Setup Setup the QC parameters that for analysis.
- System Setup Setup printer, data and language, Combo box configuration. Define sleep time and do touch screen calibration.
- Maintenance Do flow cell washing, pump calibration, optical calibration.
- About Display the version information about instrument and software.
- Power management Shut down the program or run sleep mode of instrument.
- Set the System Parameters from System Setup box (Date & Time, Printer settings etc)
- Check the functioning of printer
- Connect USB mouse or USB keypad cable to instrument USB port, check its working
- Go to maintenance and wash the flow cell two-three times with Microwash followed by distilled water.
- Go to reagent set up and create some analysis programmes like Glucose, Albumin, ALT, Creat etc.





IEi'CEi'DE'8C7 I A9BHG :CF"=BBC@56'&\$\$

ÖOCKŰÓÚŚWŰUŰÚŰW}}[læàG€€ Ű^|^æ•^\Öæc^kK€JÉ€HÉG€GG Ü^çi•i[}K€€ Ü^ci•i[}kÖæc^KÞŒ

Performance Qualification (PQ)

InnoLab200

Go to Test Menu and select any programme parameter to be tested

Run some (few enzymes, few substrate) tests using customer's reagent and fill the data in the table below

Test	QC Range (Level 1)	Test Result	QC Range (Level 2)	Test Result
GLUCOSE	84.1 (74.8-93.3)	89	271 (211-301)	255
CHOL	269 (229-309)	258	106 (90-122)	98.5
ALT/GPT	28.0(19.0-376)	26.7	104(81.0-127)	95.5
AST/GOT	44 (33.0-55.0)	41.8	193 (143-243)	189
CREAT	2.64 (2.24-3.68)	2.42	5.86 (4.98-674)	5.15
UREA	16.0 (13.6-18.4)	18	46.0(39.1-529)	41.8

	-	
D	Dar	narks
	L/GI	Hains

We hereby certify that Instrument performance check has been carried out. Verified the instrument linearity, precision & accuracy by evaluating controls. Obtained results are satisfactory. Instrument is ready for regular use.

Technical acceptance

Date (dd-mm-yyyy) :....

Service Engineer / Application Person Name

Sanjib Chattopadhyay

Signature :...

Customer Name

Signature & Stamp :