

JULY 2023 / SEROLOGY

Test method employed for detection C-reactive protein (CRP) at your lab: Latex Agglutination

Peer group (n) = 248

Please refer the attached evaluation format/answer template for details on the criteria for evaluation.

	Parameter	Your Result	Your Value (mg/L)	Intended Result	Robust Mean	Robust SD	Range (mg/L)	Uncertainty of Assigned value	Z & Z' score	Max Marks	Your Score
SE1	CRP	Negative	<6	Negative	Not Applicable					2	2
SE2	CRP	Negative	<6	Negative						2	2
SE3	CRP	Positive	6	Positive						28.6615	17.1608

Note: CRP (Latex Agglutination) peer group results of SE1 & SE2 evaluation based on interpretation because the Robust Analysis result is cannot be calculated

Disclaimer:

This is a confidential document and subject to the rules of confidentiality as described by the ISO 17043:2010 standard.

MEMBER ID:

M 0 2 5 3


SM1	SM2	SM3	CU1	CU2	CU3	SE1	SE2	SE3	Marks obtained	
3	4	4	19	19	21	2	2	2	76	100%
3	4	4	19	19	21	2	2	2	Maximum marks =76	



Dr. Rani Diana Sahni
Scientific Co-ordinator



Dr. John A Jude Prakash
Quality Manager



Dr. V. Balaji
PT Co-ordinator

Report Dispatch Date: 15.10.23

***** End of Report *****



Indian Association of Medical Microbiologists
Virology External Quality Assessment Scheme
CMCVIROEQAS, Proficiency Testing Unit, Department of Clinical Virology
Christian Medical College, Vellore-632004, Tamil Nadu
Email: viroeqas@cmcvellore.ac.in Phone: 08438172355

PC - 1034

21st SEROLOGY CMCVIREQAS EVALUATION FINAL REPORT

PANEL: BBVS

CMCVIREQAS ID. V0192

Opening Date: 18-09-2023

Closing Date: 09-10-2023

Distribution No: S323

Result receiving Date: 30.09.2023

Specimen #	Intended Result			Your Result		
	HIV	HBsAg	HCV	HIV	HBsAg	HCV
S01323009	NEGATIVE	NEGATIVE	NEGATIVE	NEGATIVE	NEGATIVE	NEGATIVE
S01323010	POSITIVE	NEGATIVE	POSITIVE	POSITIVE	NEGATIVE	POSITIVE
S01323011	NEGATIVE	POSITIVE	NEGATIVE	NEGATIVE	POSITIVE	NEGATIVE
S01323012	NEGATIVE	NEGATIVE	NEGATIVE	NEGATIVE	NEGATIVE	NEGATIVE
Your Score				8/8 (100%)	8/8 (100%)	8/8 (100%)
Performance grading				E	E	E

Scoring System

Qualitative Results	Score
Concordant Result	2
Indeterminate/Equivocal	1
Discordant Result	0



PC - 1034

**Indian Association of Medical Microbiologists
Virology External Quality Assessment Scheme
CMCVIROEQAS, Proficiency Testing Unit, Department of Clinical Virology
Christian Medical College, Vellore-632004, Tamil Nadu
Email: viroeqas@cmcvellore.ac.in Phone: 08438172355**

Assigned Value:

Assigned value is pre-determined using multiple assays/multiple testing of the same analyte.
Assigned value was not derived using reference material or calibrators.

Scoring:

Participants are scored based on qualitative result. If more than 30% of the laboratory report discrepant result that sample will not be considered for analysis.

Confidentiality of the results:

The results are kept confidential between the participant and the provider. The results can be revealed to a regulatory body with written consent from the participant. However, in exceptional circumstances, results from a particular participant will be provided to the regulatory body and the participant will be notified of this action in writing.

Appeal:

For appeals, regarding result analysis and evaluation process, please contact CMCVIROEQAS coordinator by phone 08438172355 or email viroeqas@cmcvellore.ac.in **within a week of receiving this report.**

Enquiries:

For all communications, please use your **CMCVIROEQAS LAB ID** and **Distribution Number**.

END OF REPORT

Name of CMCVIROEQAS Coordinator


Signature

Dr. Rajesh Kannangai

Department of Clinical Virology

Report Dispatch Date: 15-11-2023

Report authorized by: CMCVIROEQAS Coordinator