



CALIBRATION CERTIFICATE

HOSPITAL NAME ESIC MEDICAL COLLEGE HOSPITAL,
ADDRESS SANATHNAGAR.
ANALYSER COBAS c311
SERIAL NO 16E9-14
INSTALLATION PLACE BIOCHEMISTRY LABORATORY
DATE 04-02-2023

This is to certify that the above mentioned analyzer was calibrated by the following procedure.

1. Checked the supply voltage, and mechanical movements.
2. Adjusted the Sample Pipettor probe, Reagent Probe, Reagent Loader, and Mixer mechanism
3. Cell Blank, Photometer Check.
4. Temperature monitor for incubator, and reagent disk
5. Pipettor's probe LLD voltage, and Pressure sensor check

All the checks and calibration were done, is o.k and system is being used under recommended conditions. Next calibration is due on 03-02-2024.

For Roche Diagnostics India Pvt. Ltd.

T R Siddardha,
Sr. Technical Service Specialist,
Hyderabad.