



CERTIFICATE OF CALIBRATION/ VERIFICATION

Customer Name : E.S.I.C MEDICAL COLLEGE HOSPITAL
Model : EASYLYTE PLUS (Na/K/Cl)
Serial Number : 66437CNKC
Calibration Date : 08-03-2023
Next Calibration due on before:07-03-2024

This is to certify that the above mentioned instrument has been verified of calibration according to the procedures provided by the manufacturer in the reference manual.

The reference instruments used for value assignment are managed by the traceability systems with the manufacturer and these are traceable to international standards.

Calibration/ Verification done by

Kalyan Haldar

Kalyan Haldar
Sr. Application Manager
Technical Service Department
Transasia Bio-Medical Ltd
Hyderabad.

1. Inspection Report Enclosed



INSPECTION REPORT

Model: EASYLYTE PLUS
Serial No: 66437CNKC
Calibration Date: 08-03-2023
Quality control Used: MEDICA

General Calibration Parameters:

| Parameter | Measured Value | Acceptable Range | |
|-----------|----------------|------------------|------|
| Na+ | 57.49 | 50-64 | Pass |
| K+ | 52.90 | 50-64 | Pass |
| Cl- | 48.13 | 40-64 | Pass |

Electrode Range:

| Parameter | Electrode Value | Acceptable Range | |
|-----------|-----------------|------------------|------|
| Na+ | 125 | 100-150 | Pass |
| K+ | 75 | 50-100 | Pass |
| Cl- | 97.7 | 60-100 | Pass |

Kalyanbabu

Technical Service Department
Transasia Bio-medical Limited