

Exceptions

✓ None at this time.

Legend: ✓ No Warnings 📄 Missing Result 📅 Late Results ⚠ $2.0 \leq |Z\text{-score}| < 3.0$ ✖ $|Z\text{-score}| \geq 3.0$ * Amended Result (per participant's request)

EQAS News

IMPORTANT: Place your order for the next cycle now by contacting your Bio-Rad Representative.



Bio-Rad Laboratories is accredited to ISO/IEC 17043:2010.
Refer to the User Guide for additional information on the EQAS Program,
EQAS reports, their interpretation, statistics, and contact details.

Customer Information

Lab: 235423

**Attn: DR. IMRAN AHMED SIDDIQUI
ESIC SUPER SPECIALTY HOSPITAL
SANATH NAGAR
HYDERABAD HYDERABAD 500038
INDIA**

**Lab 235423**ESIC SUPER SPECIALTY HOSPITAL
SANATH NAGAR
HYDERABAD HYDERABAD 500038
INDIA**Configuration Report
Hemoglobin Program****Cycle 20**Dec 2022 – Nov 2023
Sample No: 11
Sample Date: 18 Oct 23
Lot No: 242000**Selected Options:**Data Input: Online
Report Receipt: QCNet
Subgroup Number: None**Instrument**

2749 – Bio–Rad D–10

Analyte	Method	Reagent	Unit
289 A1C–N	022 HPLC	0693 Bio–Rad D–10 Dual A1c (220–0201)	93 %
282 A2	022 HPLC	0664 Bio–Rad D–10 Dual HbA1c/A2/F (220–0201)	93 %



Lab 235423

ESIC SUPER SPECIALTY HOSPITAL
SANATH NAGAR
HYDERABAD HYDERABAD 500038
INDIA

Sample Summary Report
Hemoglobin Program

Cycle 20

Dec 2022 – Nov 2023
Sample No: 11
Sample Date: 18 Oct 23
Lot No: 242000



Instrument: Bio-Rad D-10

Analyte	Unit	Result	Mean	Z-score	RMZ	Comparator
✓ Hemoglobin A1c (NGSP)	%	9.5	9.30	1.16	0.11	Peer
✓ Hemoglobin A2	%	3.4	3.41	-0.07	-0.15	Peer

Legend: ✓ No Warnings 📄 Missing Result ⓘ Late Results ⚠ 2.0 ≤ |Z-score| < 3.0 ✗ |Z-score| ≥ 3.0
 * Amended Result (per participant's request) ✖ Non-robust determination of Mean and SD

Problem Classification: _____

Corrective Action: _____

Reviewed by: _____

Date: _____



Lab 235423

ESIC SUPER SPECIALTY HOSPITAL
SANATH NAGAR
HYDERABAD HYDERABAD 500038
INDIA

Hemoglobin A1c (NGSP) Report Hemoglobin Program

Cycle 20

Dec 2022 – Nov 2023
Sample No: 11
Sample Date: 18 Oct 23
Lot No: 242000



Bio-Rad D-10

Your Result






9.5 %

- ◆ Your Mode
- Your Method
- Your Peer

Comparative Statistics

	N	Mean	SD	CV	U ¹	Z-score	RMZ	%
Your Mode	1592	9.17	0.176	1.92	0.011	1.88	0.70	3.60
Your Method	1392	9.17	0.176	1.92	0.012	1.88	0.71	3.62
Your Peer	70	9.30	0.172	1.85	0.051	1.16	0.11	2.14

Your Deviation



	N	Mean	CV	U ¹
Bio-Rad D-10 HPLC				
Bio-Rad D-10 HbA1c (220-0101)/12000949	625	9.23	1.54	0.014
Bio-Rad VARIANT II TURBO/Link HPLC				
Bio-Rad TURBO 2.0 A1c Kit (12000447 / 2702455...)	180	9.13	1.77	0.030
Tosoh G7/G8/G11 Series HPLC				
Dedicated Reagent	152	8.79	1.79	0.032
Roche cobas 6000/8000/c 311				
Immunoturbidimetric Dedicated Reagent	114	9.09	2.70	0.057
Bio-Rad D-10 HPLC				
Bio-Rad D-10 Dual HbA1c/A2/F (220-0201)	82	9.23	1.73	0.044
Bio-Rad D-100 HPLC				
Bio-Rad D-100 A1c	78	9.21	1.55	0.040
Bio-Rad D-10 HPLC				
Bio-Rad D-10 Dual A1c (220-0201)	70	9.30	1.85	0.051
Sebia CAPILLARY Series				
Electrophoresis Dedicated Reagent	69	8.95	1.32	0.035
Arkray/Menarini Adams HA-8180/8190 HPLC				
Dedicated Reagent	51	8.46	2.03	0.060

Hemoglobin A1c (NGSP) 20 : 11



Lab 235423

ESIC SUPER SPECIALTY HOSPITAL
SANATH NAGAR
HYDERABAD HYDERABAD 500038
INDIA

Hemoglobin A2 Report Hemoglobin Program

Cycle 20

Dec 2022 – Nov 2023
Sample No: 11
Sample Date: 18 Oct 23
Lot No: 242000



Bio-Rad D-10






Your Result
3.4 %

- ◆ All Results
- Your Method
- Your Peer

Comparative Statistics

	N	Mean	SD	CV	U ¹	Z-score	RMZ	%
All Results	441	3.41	0.173	5.07	0.021	-0.03	-0.20	-0.168
Your Method	395	3.41	0.173	5.09	0.022	-0.06	-0.27	-0.283
Your Peer	145	3.41	0.168	4.92	0.035	-0.07	-0.15	-0.336

Your Deviation



	N	Mean	CV	U ¹
Bio-Rad VARIANT II HPLC				
Bio-Rad VARIANT II Beta-Thal Short (270-2103/...	176	3.37	4.40	0.028
Bio-Rad D-10 HPLC				
Bio-Rad D-10 Dual HbA1c/A2/F (220-0201)	145	3.41	4.92	0.035
Sebia CAPILLARY Series Electrophoresis Dedicated Reagent	42	3.37	4.99	0.065
Bio-Rad VARIANT II HPLC				
Bio-Rad VARIANT II Dual HbA1c/A2/F (270-2500)	35	3.96	7.85	0.131
Tosoh G7/G8/G11 Series HPLC Dedicated Reagent	26	3.38	6.85	0.114
Arkray/Menarini Adams HA-8180/8190 HPLC Dedicated Reagent	5	4.22	5.90	0.139
Lifetronic H9 (Triviron NANO H110)... HPLC Dedicated Reagent	4	3.53	5.10	0.225
Sebia Electrophoresis Dedicated Reagent	4	3.43	2.62	0.113
Primus Series HPLC Dedicated Reagent	1	3.50	0	0

Hemoglobin A2 20 : 11
CYCLE
SAMPLE



Lab 235423

ESIC SUPER SPECIALTY HOSPITAL
SANATH NAGAR
HYDERABAD HYDERABAD 500038
INDIA

Data on File Report
Hemoglobin Program

Cycle 20

Dec 2022 – Nov 2023
Sample No: 11
Sample Date: 18 Oct 23
Lot No: 242000



Analyte Unit	Sample Numbers											
	1	2	3	4	5	6	7	8	9	10	11	12

Instrument: Bio-Rad D-10

A1C-N	%	10.2	6.5	9.4	5.5	7.3	9.9	5.4	7.1	9.2	5.2	9.5
A2	%	2.4	3	3.5	2.5	3	2.1	2.4	2.9	3.5	2.3	3.4

Legend: No Warnings Late Results $2.0 \leq |Z\text{-score}| < 3.0$ $|Z\text{-score}| \geq 3.0$ - No Result

* Amended Result (per participant's request)