

CERTIFICATE OF CALIBRATION

CUSTOMER NAME	MODERN DIAGNOSTICS		
ADDRESS	N.S.S. BUILDING , PUTHENVELIKKARA , ERNAKULAM , KERALA 683594	DATE OF CALIBRATION	06/11/2023
CERTIFICATE NO :	CHPL/PC/MD/003	VALID UP TO	05/05/2024
		DATE OF ISSUE	07/11/2023

INSTRUMENT DETAILS			
Instrument Name	MICRO PIPETTE	S.No.	YE19BAM0056596
Make / Model No.	DRAGON LAB	Visual Inspection	OK
LOCATION :	LAB	RANGE	100 – 1000 µL

DETAILS OF REFERENCE STANDARDS & MAJOR EQUIPMENTS USED			
Equipment Name	ANALYTICAL BALANCE	THERMO HYGROMETER	
Make	WENSAR	HTC-2/-	
Model / SR No.	MAB182/39736	
Calibration Validity	27/07/2024	20/07/2024	
Calibration By	UNIQUE	UNIQUE	

ENVIRONMENTAL CONDITION			
Temperature	23.0 °C	Relative Humidity	43 %

Calibration Results							
SL NO	SET READING (µl)	MEASURED READING (µl)					STATUS
		I	II	III	IV	V	
1	100	100.00	100.01	100.02	99.99	99.98	PASS
2	500	500.00	500.01	500.02	499.99	499.98	PASS
3	1000	1000.00	1000.01	1000.02	999.99	999.98	PASS

REMARKS:-

- ❖ This result of calibration refers to only to the particular items submitted for calibration.
- ❖ The above result is valid at the time of and under the stated conditions measurement.
- ❖ Calibration certificate issued for weight & measure parameters like mass, weighting balance, volumetric equipment, measuring scales/tapes etc. Are scientific purpose only and should not used for commercial.

Calibrated By
(CALIBRATION ENGINEER)



Approved By
(TECHNICAL MANAGER)
PREJITH B

CERTIFICATE OF CALIBRATION

CUSTOMER NAME	MODERN DIAGNOSTICS		
ADDRESS	N.S.S. BUILDING , PUTHENVELIKKARA , ERNAKULAM , KERALA 683594	DATE OF CALIBRATION	06/11/2023
CERTIFICATE NO :	CHPL/PC/MD/002	VALID UP TO	05/05/2024
		DATE OF ISSUE	07/11/2023

INSTRUMENT DETAILS			
Instrument Name	MICRO PIPETTE	S.No.	YE21AAV0079222
Make / Model No.	DRAGON LAB	Visual Inspection	OK
LOCATION :	LAB	RANGE	5 – 50 µL

DETAILS OF REFERENCE STANDARDS & MAJOR EQUIPMENTS USED			
Equipment Name	ANALYTICAL BALANCE	THERMO HYGROMETER	
Make	WENSAR	HTC-2/-	
Model / SR No.	MAB182/39736	
Calibration Validity	27/07/2024	20/07/2024	
Calibration By	UNIQUE	UNIQUE	

ENVIRONMENTAL CONDITION			
Temperature	23.0 °C	Relative Humidity	43 %

Calibration Results							
SL NO	SET READING (µl)	MEASURED READING (µl)					STATUS
		I	II	III	IV	V	
1	5	5.00	5.01	4.99	4.98	4.97	PASS
2	20	20.00	20.01	20.02	19.99	19.98	PASS
3	50	50.00	50.01	50.02	49.99	49.98	PASS

REMARKS:-

- ❖ This result of calibration refers to only to the particular items submitted for calibration.
- ❖ The above result is valid at the time of and under the stated conditions measurement.
- ❖ Calibration certificate issued for weight & measure parameters like mass, weighting balance, volumetric equipment, measuring scales/tapes etc. Are scientific purpose only and should not used for commercial.

Calibrated By
(CALIBRATION ENGINEER)



Approved By
(TECHNICAL MANAGER)
PREJITH T B

CERTIFICATE OF CALIBRATION

CUSTOMER NAME	MODERN DIAGNOSTICS		
ADDRESS	N.S.S. BUILDING , PUTHENVELIKKARA , ERNAKULAM , KERALA 683594	DATE OF CALIBRATION	06/11/2023
CERTIFICATE NO :	CHPL/PC/MD/001	VALID UP TO	05/05/2024
		DATE OF ISSUE	07/11/2023

INSTRUMENT DETAILS			
Instrument Name	MICRO PIPETTE	S.No.	YE5A429495
Make / Model No.	DRAGON LAB	Visual Inspection	OK
LOCATION :	LAB	RANGE	100 µL

DETAILS OF REFERENCE STANDARDS & MAJOR EQUIPMENTS USED			
Equipment Name	ANALYTICAL BALANCE	THERMO HYGROMETER	
Make	WENSAR	HTC-2/-	
Model / SR No.	MAB182/39736	
Calibration Validity	27/07/2024	20/07/2024	
Calibration By	UNIQUE	UNIQUE	

ENVIRONMENTAL CONDITION			
Temperature	23.0 °C	Relative Humidity	43 %

Calibration Results							
SL NO	SET READING (µl)	MEASURED READING (µl)					STATUS
		I	II	III	IV	V	
1	100	100.00	100.01	100.02	99.99	99.98	PASS

REMARKS:-

- ❖ This result of calibration refers to only to the particular items submitted for calibration.
- ❖ The above result is valid at the time of and under the stated conditions measurement.
- ❖ Calibration certificate issued for weight & measure parameters like mass, weighting balance, volumetric equipment, measuring scales/tapes etc. Are scientific purpose only and should not used for commercial.

Calibrated By
(CALIBRATION ENGINEER)



Approved By
(TECHNICAL MANAGER)
PREJITH T B


