

CALIBRATION CERTIFICATE

NAME OF THE CUSTOMER ESIC Hospital Sanath Nagar, Hyderabad, Telangana – 500038.	MAKE Sysmex Corporation Sysmex India Pvt. Ltd. 1002, Damji Shamji Business Galleria, 10th Floor, LBS Marg Kanjurmarg (West), Mumbai 400 078, India Tel: +91 (22) 6112 6666 Fax: +91 (22) 2577 6790
---	---

PRODUCT NAME XN-1000 HEAMATOLOGY ANALYZER	SERIAL NO 47607
--	-------------------------------

This is to certify that the XN-1000 Hematology Analyzer has been calibrated with XN CAL Calibrator (Lot No. 23322101, Exp. Date 01/01/2023)

S.NO	TEST PARAMETERS	MEASURED/CALIBRATED	REMARKS
1.	Start up	WBC/RBC/HGB/PLT	PASS
2.	Reproducibility	WBC/RBC/HGB/MCV/HCT/PLT	PASS
3.	QC	WBC/RBC/HGB/HCT/PLT	PASS
4.	Calibration (Lot No. 23322101)	WBC/RBC/HGB/HCT/PLT	PASS

Date of Calibration: 10/11/2022	Due date for Calibration: 09/12/2023
---------------------------------	--------------------------------------

NAME OF THE ENGINEER

SIGNATURE

Sridhar Dasaradhi

