

BIO-RAD LABORATORIES (INDIA) PRIVATE LIMITED

Site name:	ESIC SUPER SPECIALITY HOSPITAL		
Address:	HYDERABAD		
Province-Postal code	500038		
Contact Name	Dr.Imran		

Calibration Certificate

Date	27-06-2023	Instrument	D 10 Ub Apolyzor	Instrument	DJ8E226202
	(dd mmmm, yyyy)	Model	D-10 Hb Analyzer	Serial Number	D30E220202

We certify that the Bio-Rad instrument specified above and installed in your lab is calibrated and the performance was found satisfactory.

High Pressure Pump Flow Rate Reference: Service Manual: Section 7.7.3

Measured with (Device manufacturer and model	Device Serial Number	Device Calibration date (dd mmmm, yyyy)
GLASSCO MC	2118	20-11-2021

Specified Range@ 1.5ml/Minute	Measured Value
1.425 - 1.575 ml/minute	1.50ML/mINUTE

Column Holder temperature Reference: Service Manual: Section 7.9

Measured with (Device manufacturer and model	Device Serial Number	Device Calibration date (dd mmmm, yyyy)
Fluke -179	31090109	31-05-2023
Specified Range @ 35°C	Measured Value	
33.4°C – 36.4°C	35.1	

Detector log value Reference: Service Manual: Section 7.2 or Field Bulletin D-10-081

Measured with	Specified Range	Measured Value
Instrument Service Software	0.018 – 0.022 AU	0.020AU

The next calibration is due on	26-06-2024	(dd mmmm, vvvv)
--------------------------------	------------	-----------------

For Bio-Rad Laboratories India Pvt Limited

Authorized Signatory:

GRas.

Page 1 of 1 Calibration Certificate