



Calibrator Calibration

Instrument: 5 Part Haematology Analyzer		Model: BC -6200		Serial No.: TW-04001033		Date of Cal. 06.11.2023
Mode: OV-WB		Calibrator Lot No. PLUS0923			Exp.Date:05.10.2023	
Parameter	WBC	RBC	HGB	MCV	PLT	WBC-D
TARGET	8.98	4.61	13.3	91.1	254	8.98
1.	8.71	4.77	14.2	91.9	250	8.58
2.	8.69	4.74	14.0	92.0	254	8.51
3.	8.67	4.73	14.1	91.7	261	8.52
4.	8.95	4.79	14.0	91.6	251	8.53
5.	9.09	4.77	14.1	91.8	250	8.64
6.	8.65	4.76	14.1	91.9	251	8.91
Mean	8.79	4.76	14.1	91.8	253	8.61
CV(%)	2.072	0.436	0.44	0.17	1.7	1.764
New Cal. Factor (%)	103.21	98.64	96.12	106.02	101.62	
Old Cal. Factors (%)	92.65	89.86	96.12	106.02	103.12	
Old TC1 (%)						100.84
New TC1 (%)						102.10

For Rapid Diagnostic Pvt. Ltd.

B. NARASIMHA
Mahender Reddy
 Service Engineer



N.

Zonal Service Manager



Calibration Certificate

Customer Name	Telangana Diagnostics ASIFABAD
Address	ESI HOSPITAL, HYDRABAD-504001
Contact Person	GEETHA
Contact No.	9440625010
Instrument	5 Part Haematology Analyzer
Model	BC 6200
Serial No.	TW-03000931
Instrument Status	Under Warranty
Calibration Done On	06.11.2023
Next Calibration Due On	06.11.2024

For Rapid Diagnostic Pvt. Ltd.

B. NARASIMHA

N. Mahender Reddy
Service Engineer

Zonal Service Manager



Rapid Diagnostic Pvt. Ltd.

