



AACPL CALIBRATION PVT. LTD.

CALIBRATION CERTIFICATE

7.8F-01



CC - 2646

Certificate No	AACPL / 06975F	Field	Mech.	Group	Volume	Page No.	1 of 1
SRF No & Date	230918.2	DT - 18 / 09 / 2023	ULR No.		CC2646230000 06975F		
Calibration Method :	Comparison Method			Accuracy			
Company Name	M/s Krsnaa Diagnostics Ltd.			Calibration Date		18.09.2023	
Address	SDH, Civil Hospital Road, Shastri Nagar, Batala, Punjab			Recommended Date for Next Calib.		17.09.2024	
				Certificate Issue Dt.		20.09.2023	
				UUC Condition		Good	

Calibration Instrument Detail

Sl	Instrument Name	Range	Least Count	Make / SI.No.	Party ID / Location
1	Micro Pipette	10 to 1000 µl	5 µl	Dragon Lab YE225AX0313175	KDSDPBCCMP10 Bio Chemistry

Standard Equipments Used (Traceable to National Standard)

Sl	Instrument Name	Make/SI No	Calibrated By	Cal Certificate No	Cal. Valid Upto
1	Weighing Balance	Radwag / 668683	Bluebox Techno	BBT/452/JAN/23	05/03/2024

Reference Standard	Calibration Purpose	Humidity	Temperature	Calibration Performed At
ISO - 8655 - 6	Volume Measurement	(50 ± 10)%RH	(23 ± 2)°C	Laboratory

Calibration Procedure : As per Calibration Procedure Manual AACPL / CP / 01C, Section No : CP - 03

Calibration Result

S. No	Nominal Value on UUC (in µl)	Average Volume at 27 °C (in µl)	Deviation (in µl)	Uncertainty at approx 95% confidence Level and coverage factor $k = 2$ is ± (in µl)
1	100	100.00256	-0.00256	0.25
2	500	500.01431	-0.01431	0.25
3	1000	1000.16874	-0.16874	0.25

Note

- 1 The calibration results reported in this calibration certificate are valid at the time of & under stated condition
- 2 This certificate cannot be reproduced except in full without our prior permission in writing
- 3 This certificate refers only to the particular item(s) submitted for calibration
- 4 UUC - Unit under calibration

----- End of Certificate -----

Astidh
Calibrated by

Approved by
[Signature]
Anand Chaurasia
(CEO)



AACPL CALIBRATION PVT. LTD.

CALIBRATION CERTIFICATE

7.8F-01



CC - 2646

Certificate No	AACPL / 06974F	Field	Mech.	Group	Volume	Page No.	1 of 1
SRF No & Date	230918.2	DT - 18 / 09 / 2023		ULR No.	CC2646230000 06974F		
Calibration Method :	Comparison Method			Accuracy			
Company Name	M/s Krsnaa Diagnostics Ltd.			Calibration Date	18.09.2023		
Address	SDH, Civil Hospital Road, Shastri Nagar, Batala, Punjab			Recommended Date for Next Calib.	17.09.2024		
				Certificate Issue Dt.	20.09.2023		
				UUC Condition	Good		

Calibration Instrument Detail

Sl	Instrument Name	Range	Least Count	Make / SI.No.	Party ID / Location
1	Micro Pipette	20 to 200 µl	0.1 µl	Micropiet / ---	KDSDPBCCMP09 Bio Chemistry

Standard Equipments Used (Traceable to National Standard)

Sl	Instrument Name	Make/SI No	Calibrated By	Cal Certificate No	Cal. Valid Upto
1	Weighing Balance	Radwag / 668683	Bluebox Techno	BBT/452/JAN/23	05/03/2024

Reference Standard	Calibration Purpose	Humidity	Temperature	Calibration Performed At
ISO - 8655 - 6	Volume Measurement	(50 ± 10)%RH	(23 ± 2)°C	Laboratory

Calibration Procedure : As per Calibration Procedure Manual AACPL / CP / 01C, Section No : CP - 03

Calibration Result

S. No	Nominal Value on UUC (in µl)	Average Volume at 27 °C (in µl)	Deviation (in µl)	Uncertainty at approx 95% confidence Level and coverage factor $k = 2$ is ± (in µl)
1	20	20.09711	-0.09711	0.25
2	100	100.31569	-0.31569	0.25
3	200	200.74695	-0.74695	0.25

Note

- The calibration results reported in this calibration certificate are valid at the time of & under stated condition
- This certificate cannot be reproduced except in full without our prior permission in writing
- This certificate refers only to the particular item(s) submitted for calibration
- UUC - Unit under calibration

----- End of Certificate -----

Adish
Calibrated by



Approved by
Anand Chaurasia
Anand Chaurasia
(CEO)



AAC CALIBRATION PVT. LTD.

CALIBRATION CERTIFICATE

7.8F-01



CC - 2646

Certificate No	AACPL / 06971F	Field	Elec.Tech.	Group		Page No.	1 of 1
SRF No & Date	230918.1	DT - 18 / 09 / 2023		ULR No.	CC2646230000	06971F	
Calibration Method :	Direct Method			Accuracy			
Company Name	M/s Krsnaa Diagnostics Ltd.			Calibration Date	18.09.2023		
Address	SDH, Civil Hospital Road, Shastri Nagar, Batala, Punjab			Recommended Date for Next Calib.	17.09.2024		
				Certificate Issue Dt.	20.09.2023		
				UUC Condition	Good		

Calibration Instrument Detail

SI	Instrument Name	Range	Least Count	Make / SI.No.	Party ID / Location
1	Dig. Temp. Controller	2 to 8 °C	1 °C	----	KDSDPBSTD17 Store Room (Refrigerator)

Standard Equipments Used (Traceable to National Standard)

SI	Instrument Name	Make/SI No	Calibrated By	Cal Certificate No	Cal. Valid Upto
1	Universal Calibrator	Radix / 112004873	AACPL	AACPL/00047F	03/04/2024

Reference Standard	Calibration Purpose	Humidity	Temperature	Calibration Performed At
IS : CG-11	Temp. Measurement	(50 ± 20)%RH	(25 ± 10)°C	Site

Calibration Procedure : As per Calibration Procedure Manual AACPL / CP / 03, Section No : CPET - 01

Set Points :- 2, 5 & 8 °C

Calibration Results

SI. No.	Std. Reading on Master in °C	UUC Average Reading in °C	Deviation in °C	Uncertainty at approx 95% confidence Level and coverage factor k =2 is ±
1	2.1	3	0.9	1.62 °C
2	5.1	6	0.9	1.62 °C
3	8.1	9	0.9	1.62 °C

Note

- 1 The calibration results reported in this calibration certificate are valid at the time of & under stated condition
- 2 This certificate cannot be reproduced except in full without our prior permission in writing
- 3 This certificate refers only to the particular item(s) submitted for calibration
- 4 UUC - Unit under calibration

----- End of Certificate -----

Calibrated by



Approved by
Bhoopesh Kumar
(Tech. Manager)



A A CALIBRATION PVT. LTD.

CALIBRATION CERTIFICATE

7.8F-01



CC - 2646

Certificate No	AACPL / 06969F	Field	Elec.Tech.	Group		Page No.	1 of 1
SRF No & Date	230918.1	DT - 18 / 09 / 2023		ULR No.	CC2646230000	06969F	
Calibration Method :	Direct Method			Accuracy			
Company Name	M/s Krsnaa Diagnostics Ltd.			Calibration Date	18.09.2023		
Address	SDH, Civil Hospital Road, Shastri Nagar, Batala, Punjab			Recommended Date for Next Calib.	17.09.2024		
				Certificate Issue Dt.	20.09.2023		
				UUC Condition	Good		

Calibration Instrument Detail

SI	Instrument Name	Range	Least Count	Make / SI.No.	Party ID / Location
1	Dig. Temp. Controller	2 to 8 °C	1 °C	Samsung RT28A3022GS/NL/21	KDSDPBSTRF15 Store Room (Refrigerator)

Standard Equipments Used (Traceable to National Standard)

SI	Instrument Name	Make/SI No	Calibrated By	Cal Certificate No	Cal. Valid Upto
1	Universal Calibrator	Radix / 112004873	AACPL	AACPL/00047F	03/04/2024

Reference Standard	Calibration Purpose	Humidity	Temperature	Calibration Performed At
IS : CG-11	Temp. Measurement	(50 ± 20)%RH	(25 ± 10)°C	Site

Calibration Procedure : As per Calibration Procedure Manual AACPL / CP / 03, Section No : CPET - 01

Set Points :- 2, 5 & 8 °C

Calibration Results

SI. No.	Std. Reading on Master in °C	UUC Average Reading in °C	Deviation in °C	Uncertainty at approx 95% confidence Level and coverage factor k =2 is ±
1	2.1	2	-0.1	1.62 °C
2	5.1	5	-0.1	1.62 °C
3	8.1	9	0.9	1.62 °C

Note

- 1 The calibration results reported in this calibration certificate are valid at the time of & under stated condition
- 2 This certificate cannot be reproduced except in full without our prior permission in writing
- 3 This certificate refers only to the particular item(s) submitted for calibration
- 4 UUC - Unit under calibration

----- End of Certificate -----

Calibrated by



Approved by

Bhoopesh Kumar
(Tech. Manager)



AAC CALIBRATION PVT. LTD.

CALIBRATION CERTIFICATE

7.8F-01



CC - 2648

Certificate No	AACPL / 06968F	Field	Elec.Tech.	Group		Page No.	1 of 1
SRF No & Date	230918.1	DT - 18 / 09 / 2023		ULR No.	CC2646230000 06968F		
Calibration Method :	Direct Method			Accuracy			
Company Name	M/s Krsnaa Diagnostics Ltd.			Calibration Date	18.09.2023		
Address	SDH, Civil Hospital Road, Shastri Nagar, Batala, Punjab			Recommended Date for Next Calib.	17.09.2024		
				Certificate Issue Dt.	20.09.2023		
				UUC Condition	Good		

Calibration Instrument Detail

SI	Instrument Name	Range	Least Count	Make / SI.No.	Party ID / Location
1	Dig. Temp. Controller	0 to 110 °C	1 °C	ARBA Healthcare ----	KDSDPBCCIN11 Boi Chenistry (Incubator)

Standard Equipments Used (Traceable to National Standard)

SI	Instrument Name	Make/SI No	Calibrated By	Cal Certificate No	Cal. Valid Upto
1	Universal Calibrator	Radix / 112004873	AACPL	AACPL/00047F	03/04/2024

Reference Standard	Calibration Purpose	Humidity	Temperature	Calibration Performed At
IS : CG-11	Temp. Measurement	(50 ± 20)%RH	(25 ± 10)°C	Site

Calibration Procedure : As per Calibration Procedure Manual AACPL / CP / 03, Section No : CPET - 01

Set Points :- 10, 60 & 110 °C

Calibration Results

Sl. No.	Std. Reading on Master in °C	UUC Average Reading in °C	Deviation in °C	Uncertainty at approx 95% confidence Level and coverage factor k =2 is ±
1	10.1	10	-0.1	1.62 °C
2	60.1	60	-0.1	1.62 °C
3	110.2	111	0.8	1.62 °C

Note

- 1 The calibration results reported in this calibration certificate are valid at the time of & under stated condition
- 2 This certificate cannot be reproduced except in full without our prior permission in writing
- 3 This certificate refers only to the particular item(s) submitted for calibration
- 4 UUC - Unit under calibration

----- End of Certificate -----

Calibrated by



Approved by

Bhoopesh Kumar
(Tech. Manager)



A A CALIBRATION PVT. LTD.

CALIBRATION CERTIFICATE

7.8F-01



CC - 2646

Certificate No	AACPL / 06967F	Field	Elec.Tech.	Group	Page No.	1 of 1
SRF No & Date	230918.1	DT - 18 / 09 / 2023		ULR No.	CC2646230000 06967F	
Calibration Method :	Direct Method			Accuracy		
Company Name	M/s Krsnaa Diagnostics Ltd.			Calibration Date	18.09.2023	
Address	SDH, Civil Hospital Road, Shastri Nagar, Batala, Punjab			Recommended Date for Next Calib.	17.09.2024	
				Certificate Issue Dt.	20.09.2023	
				UUC Condition	Good	

Calibration Instrument Detail

SI	Instrument Name	Range	Least Count	Make / SI.No.	Party ID / Location
1	Dig. Timer	0 - 99 min	1 min	Remi ZHFN - 1968	KDSDPBCPCF07 // Clinical Pathology (Centrifuge Machine)

Standard Equipments Used (Traceable to National Standard)

SI	Instrument Name	Make/SI No	Calibrated By	Cal Certificate No	Cal. Valid Upto
1	Dig. Time Interval Calibrator	Voltagram / -----	VTC	CC293522000002078F	27/10/2023

Reference Standard	Calibration Purpose	Humidity	Temperature	Calibration performed At
NIST 960 - 12	Time Measurement	(50 ± 20) % RH	(25 ± 10)°C	Site

Calibration Procedure : As per Calibration Procedure Manual AACPL / CP / 03, Section No : CPET - 02

Calibration Result

SI	Selected Range (In s)	Average Read. (Std.) (In s)	Average Read. (UUC) (In min)	Deviation (In s)	Uncertainty at approx 95% confidence Level and coverage factor $k = 2$ is ±
1	600	599.77	10.0	0.23	34.67 s
2	900	899.75	15.0	0.25	34.67 s
3	1800	1799.2	30.0	0.8	34.67 s

Note

- 1 The calibration results reported in this calibration certificate are valid at the time of & under stated condition
- 2 This certificate cannot be reproduced except in full without our prior permission in writing
- 3 This certificate refers only to the particular item(s) submitted for calibration
- 4 UUC - Unit under calibration

----- End of Certificate -----

Calibrated by



Approved by

Bhoopesh Kumar
(Tech. Manager)



A A CALIBRATION PVT. LTD.

CALIBRATION CERTIFICATE

7.8F-01



CC - 2646

Certificate No	AACPL / 06970F	Field	Mech.	Group	Pressure	Page No.	1 of 1
SRF No & Date	230918.1	DT - 18 / 09 / 2023		ULR No.	CC2646230000 06970F		
Calibration Method :	Comparison Method			Accuracy			
Company Name	M/s Ksrnaa Diagnostics Ltd.			Calibration Date	18.09.2023		
Address	SDH, Civil Hospital Road, Shastri Nagar, Batala, Punjab			Recommended Date for Next Calib.	17.09.2024		
				Certificate Issue Dt.	20.09.2023		
				UUC Condition	Good		

Calibration Instrument Detail

SI	Instrument Name	Range	Least Count	Make / SI.No.	Party ID / Location
1	Pressure Gauge	0 - 2.1 kg/cm ²	0.05 kg/cm ²	ARBA Healthcare -----	KSDPBSTAU16 Store Room (Autoclave)

Standard Equipments Used (Traceable to National Standard)

SI	Instrument Name	Make/SI No	Calibrated By	Cal Certificate No	Cal. Valid Upto
1	Dig. Pressure Gauge	R & D Inst. / 35691904	R & D Inst. Services	CS/23/LB/MP/62-01	16/05/2024

Reference Standard	Calibration Purpose	Humidity	Temperature	Calibration Performed At
DKD-R6-1 - 2014	Pressure Measurement	(50 ± 20)%RH	(23 ± 5) °C	Site

Calibration Method : Pneumatic Comparison Method

Calibration Procedure : As per Calibration Procedure Manual AACPL / CP / 01B, Section No : CP - 01

Uncertainty Measurement at approx 95% confidence level and coverage factor $k = 2$ is ± 0.036 kg/cm²

Calibration Result (In kg/cm²)

SI	Calibration Gauge Value	Standard Gauge Value			
		Increasing	Deviation	Decreasing	Deviation
No	UUC				
1	0.00	0.0000	0.0000	0.0000	0.0000
2	0.20	0.2005	-0.0005	0.2003	-0.0003
3	0.50	0.5009	-0.0009	0.5007	-0.0007
4	1.00	1.0013	-0.0013	1.0012	-0.0012
5	1.50	1.5027	-0.0027	1.5025	-0.0025
6	2.10	2.1034	-0.0034	2.1031	-0.0031

Note

- The calibration results reported in this calibration certificate are valid at the time of & under stated condition
- This certificate cannot be reproduced except in full without our prior permission in writing
- This certificate refers only to the particular item(s) submitted for calibration
- UUC - Unit under calibration
- Master Instrument calibration certificate are traceable to SI Unit

----- End of Certificate -----


Calibrated by



Approved by

Vinit Vats
(Calibration Engineer)

LAB :- Plot No. C-18, Sector-13, First Floor, "Vasundhara, Haat" Vasundhara, Ghaziabad - 201012

Ph. : 0120-4135529, Mob. : 9212172564, 9212172597

E-mail : aacplinfo2005@gmail.com. aacplinfo@rediffmail.com



AACALIBRATION PVT. LTD.

CALIBRATION CERTIFICATE

7.8F-01



CC - 2646

Certificate No	AACPL / 06966F	Field	Mech.	Group	Speed	Page No.	1 of 1
SRF No & Date	230918.1	DT - 18 / 09 / 2023		ULR No.	CC2646230000 06966F		
Calibration Method :	Comparison Method			Accuracy			
Company Name	M/s Krsnaa Diagnostics Ltd.			Calibration Date	18.09.2023		
Address	SDH, Civil Hospital Road, Shastri Nagar, Batala, Punjab			Recommended Date for Next Calib.	17.09.2024		
				Certificate Issue Dt.	20.09.2023		
				UUC Condition	Good		

Calibration Instrument Detail

SI	Instrument Name	Range	Least Count	Make / SI.No.	Party ID / Location
1	Centrifuge Machine	0 - 6000 RPM	10 RPM	Remi ZHFN-1968	KDSDPBCPCF07 Clinical Pathology

Standard Equipments Used (Traceable to National Standard)

SI	Instrument Name	Make/SI No	Calibrated By	Cal Certificate No	Cal. Valid Upto
1	Dig. Tachometer	Line Seiki / 20120427	Godrej Calibration Services	C-230215-9-1	16/02/2024

Reference Standard	Calibration Purpose	Humidity	Temperature	Calibration Performed At
IS : 12508 - 2018	RPM Measurement	(50 ± 20)%RH	(25 ± 10) °C	Site

Calibration Procedure : As per Calibration Procedure Manual AACPL / CP / 01B, Section No : CP - 08

Calibration Result

SI	Calib. Gauge Value	Master Gauge Value	Deviation	Uncertainty at approx 95% confidence Level and coverage factor $k = 2$ is \pm (in RPM)
	UUC (in RPM)	(in RPM)	(in RPM)	
1	100	100.1	-0.1	5.88
2	500	500.5	-0.5	5.88
3	1000	1000.8	-0.8	5.88
4	2000	2001.1	-1.1	6.15
5	6000	6001.5	-1.5	6.15

Note

- 1 The calibration results reported in this calibration certificate are valid at the time of & under stated condition
- 2 This certificate cannot be reproduced except in full without our prior permission in writing
- 3 This certificate refers only to the particular item(s) submitted for calibration
- 4 UUC - Unit under calibration

----- End of Certificate -----

Calibrated by 



Approved by

Vinit Vats
(Calibration Engineer)

LAB :- Plot No. C-18, Sector-13, First Floor, "Vasundhara, Haat" Vasundhara, Ghaziabad - 201012

Ph. : 0120-4135529, Mob. : 9212172564, 9212172597

E-mail : aacplinfo2005@gmail.com. aacplinfo@rediffmail.com



AACPL CALIBRATION PVT. LTD.

CALIBRATION CERTIFICATE

7.8F-01



CC - 2646

Certificate No	AACPL / 06976F	Field	Thermal	Group	Temp. & RH	Page No.	1 of 1
SRF No & Date	230918.2	DT - 18 / 09 / 2023		ULR No.	CC2646230000 06976F		
Calibration Method :	Comparison Method			Accuracy			
Company Name	M/s Krsnaa Diagnostics Ltd.			Calibration Date	18.09.2023		
Address	SDH, Civil Hospital Road, Shastri Nagar, Batala, Punjab			Recommended Date for Next Callb.	17.09.2024		
				Certificate Issue Dt.	20.09.2023		
				UUC Condition	Good		

Calibration Instrument Detail

Sl	Instrument Name	Range	Least Count	Make / Si.No.	Party ID / Location
1	Dig. Thermo Hygrometer	-10 to 50 °C 10 to 99 % RH	0.1 °C / 1 % RH	HTC - 1 / ---	KDSDPBCCTH12 Bio Chemistry

Standard Equipments Used (Traceable to National Standard)

Sl	Instrument Name	Make/Si.No	Calibrated At	Cal Certificate No	Cal. Valid Upto
1	Dig. Hygrometer With Sensor	Polltech/20275542	NCL	NCL/A/2023/67/01	21/01/2024
2	Humidity Chamber	Indecon / -----	-----	-----	-----

Reference Standard	Calibration Purpose	Humidity	Temperature	Calibration Performed At
Comparison Method	Temp. & RH Measurement	(50 ± 15)%RH	(25 ± 2)°C	Laboratory

Calibration Procedure : As per Calibration Procedure Manual AACPL / CP / 03, Section No : CPTH - 07

Calibration Points :- Temp. 15, 25 & 40 °C Relative Humidity 30, 60 & 90 % RH

Calibration Results (Temperature at 50 % RH)

Sl. No.	Master Value on Std. (In °C)	UUC Value (In °C)	Deviation (In °C)	Uncertainty at approx 95% confidence Level and coverage factor k =2 is ± (°C)
1	15.06	14.9	-0.16	0.47
2	25.13	24.7	-0.43	0.47
3	40.17	39.6	-0.57	0.47

Calibration Results (Humidity at 25 °C)

Sl. No.	Master Value on Std. (In % RH)	UUC Value (In % RH)	Deviation (In % RH)	Uncertainty at approx 95% confidence Level and coverage factor k =2 is ± (%RH)
1	30.05	30	-0.05	1.77
2	60.09	60	-0.09	1.77
3	90.12	90	-0.12	1.77

Note

- The calibration results reported in this calibration certificate are valid at the time of & under stated condition
- This certificate cannot be reproduced except in full without our prior permission in writing
- This certificate refers only to the particular item(s) submitted for calibration
- UUC - Unit under calibration

----- End of Certificate -----

[Signature]
Calibrated by



Approved by
[Signature]
Anand Chaurasia
(CEO)

LAB :- Plot No. C-18, Sector-13, First Floor, "Vasundhara, Haat" Vasundhara, Ghaziabad - 201012

Ph. : 0120-4135529, Mob. : 9212172564, 9212172597

E-mail : aacplinfo2005@gmail.com, aacplinfo@rediffmail.com



AACPL CALIBRATION PVT. LTD.

CALIBRATION CERTIFICATE

7.8F-01



CC - 2646

Certificate No	AACPL / 06973F	Field	Mech.	Group	Volume	Page No.	1 of 1
SRF No & Date	230918.2	DT - 18 / 09 / 2023		ULR No.	CC2646230000 06973F		
Calibration Method :	Comparison Method			Accuracy			
Company Name	M/s Krsnaa Diagnostics Ltd.			Calibration Date	18.09.2023		
Address	SDH, Civil Hospital Road, Shastri Nagar, Batala, Punjab			Recommended Date for Next Calib.	17.09.2024		
				Certificate Issue Dt.	20.09.2023		
				UUC Condition	Good		

Calibration Instrument Detail

Sl	Instrument Name	Range	Least Count	Make / Sl.No.	Party ID / Location
1	Micro Pipette	100 µl	-----	Diaplast / ---	KDSDPBCCMP08 Bio Chemistry

Standard Equipments Used (Traceable to National Standard)

Sl	Instrument Name	Make/Sl No	Calibrated By	Cal Certificate No	Cal. Valid Upto
1	Weighing Balance	Radwag / 668683	Bluebox Techno	BBT/452/JAN/23	05/03/2024

Reference Standard	Calibration Purpose	Humidity	Temperature	Calibration Performed At
ISO - 8655 - 6	Volume Measurement	(50 ± 10)%RH	(23 ± 2)°C	Laboratory

Calibration Procedure : As per Calibration Procedure Manual AACPL / CP / 01C, Section No : CP - 03

Calibration Result

S. No	Nominal Value on UUC (in µl)	Average Volume at 27 °C (in µl)	Deviation (in µl)	Uncertainty at approx 95% confidence Level and coverage factor $k = 2$ is ± (in µl)
1	100	99.72453	0.27547	0.25

Note

- 1 The calibration results reported in this calibration certificate are valid at the time of & under stated condition
- 2 This certificate cannot be reproduced except in full without our prior permission in writing
- 3 This certificate refers only to the particular item(s) submitted for calibration
- 4 UUC - Unit under calibration

----- End of Certificate -----


Calibrated by



Approved by

Anand Chaurasia
(CEO)

LAB :- Plot No. C-18, Sector-13, First Floor, "Vasundhara, Haat" Vasundhara, Ghaziabad - 201012

Ph. : 0120-4135529, Mob. : 9212172564, 9212172597

E.mail : aacplinfo2005@gmail.com, aacplinfo@rediffmail.com



A A CALIBRATION PVT. LTD.

CALIBRATION CERTIFICATE

7.8F-01



CC - 2646

Certificate No	AACPL / 06978F	Field	Thermal	Group	Temp. & RH	Page No.	1 of 1
SRF No & Date	230918.2	DT - 18 / 09 / 2023		ULR No.	CC2646230000	06978F	
Calibration Method :	Comparison Method			Accuracy			
Company Name	M/s Krsnaa Diagnostics Ltd.			Calibration Date	18.09.2023		
Address	SDH, Civil Hospital Road, Shastri Nagar, Batala, Punjab			Recommended Date for Next Calib.	17.09.2024		
				Certificate Issue Dt.	20.09.2023		
				UUC Condition	Good		

Calibration Instrument Detail

SI	Instrument Name	Range	Least Count	Make / SI.No.	Party ID / Location
1	Dig. Thermo Hygrometer	-10 to 50 °C 10 to 99 % RH	0.1 °C / 1 % RH	HTC - 1 / ---	KDSDPBSCTH14 Sample Collection Room

Standard Equipments Used (Traceable to National Standard)

SI	Instrument Name	Make/SI.No	Calibrated At	Cal Certificate No	Cal. Valid Upto
1	Dig. Hygrometer With Sensor	Polltech/20275542	NCL	NCL/A/2023/67/01	21/01/2024
2	Humidity Chamber	Indecon / -----	-----	-----	-----

Reference Standard	Calibration Purpose	Humidity	Temperature	Calibration Performed At
Comparison Method	Temp. & RH Measurement	(50 ± 15)%RH	(25 ± 2)°C	Laboratory

Calibration Procedure : As per Calibration Procedure Manual AACPL / CP / 03, Section No : CPTH - 07

Calibration Points :- Temp. 15, 25 & 40 °C Relative Humidity 30, 60 & 90 % RH

Calibration Results (Temperature at 50 % RH)

SI. No.	Master Value on Std. (In °C)	UUC Value (In °C)	Deviation (In °C)	Uncertainty at approx 95% confidence Level and coverage factor $k=2$ is \pm (°C)
1	15.06	15.1	0.04	0.47
2	25.13	25.2	0.07	0.47
3	40.17	40.4	0.23	0.47

Calibration Results (Humidity at 25 °C)

SI. No.	Master Value on Std. (In % RH)	UUC Value (In % RH)	Deviation (In % RH)	Uncertainty at approx 95% confidence Level and coverage factor $k=2$ is \pm (%RH)
1	30.05	30	-0.05	1.77
2	60.09	60	-0.09	1.77
3	90.12	89	-1.12	1.77

Note

- 1 The calibration results reported in this calibration certificate are valid at the time of & under stated condition
- 2 This certificate cannot be reproduced except in full without our prior permission in writing
- 3 This certificate refers only to the particular item(s) submitted for calibration
- 4 UUC - Unit under calibration

----- End of Certificate -----

Calibrated by



Approved by

Anand Chaurasia
(CEO)

LAB :- Plot No. C-18, Sector-13, First Floor, "Vasundhara, Haat" Vasundhara, Ghaziabad - 201012

Ph. : 0120-4135529, Mob. : 9212172564, 9212172597

E-mail : aacplinfo2005@gmail.com, aacplinfo@rediffmail.com



AACPL CALIBRATION PVT. LTD.

CALIBRATION CERTIFICATE

7.8F-01



CC - 2646

Certificate No	AACPL / 06977F	Field	Thermal	Group	Temp. & RH	Page No.	1 of 1
SRF No & Date	230918.2	DT - 18 / 09 / 2023		ULR No.	CC2646230000 06977F		
Calibration Method :	Comparison Method			Accuracy			
Company Name	M/s Krsnaa Diagnostics Ltd.			Calibration Date	18.09.2023		
Address	SDH, Civil Hospital Road, Shastri Nagar, Batala, Punjab			Recommended Date for Next Calib.	17.09.2024		
				Certificate Issue Dt.	20.09.2023		
				UUC Condition	Good		

Calibration Instrument Detail

SI	Instrument Name	Range	Least Count	Make / SI.No.	Party ID / Location
1	Dig. Thermo Hygrometer	-10 to 50 °C 10 to 99 % RH	0.1 °C / 1 % RH	HTC - 1 / ---	KDSDPBHMT13 Haematology

Standard Equipments Used (Traceable to National Standard)

SI	Instrument Name	Make/SI.No	Calibrated At	Cal Certificate No	Cal. Valid Upto
1	Dig. Hygrometer With Sensor	Polltech/20275542	NCL	NCL/A/2023/67/01	21/01/2024
2	Humidity Chamber	Indecon / -----	-----	-----	-----

Reference Standard	Calibration Purpose	Humidity	Temperature	Calibration Performed At
Comparison Method	Temp. & RH Measurement	(50 ± 15)%RH	(25 ± 2)°C	Laboratory

Calibration Procedure : As per Calibration Procedure Manual AACPL / CP / 03, Section No : CPTH - 07

Calibration Points :- Temp. 15, 25 & 40 °C Relative Humidity 30, 60 & 90 % RH

Calibration Results (Temperature at 50 % RH)

SI. No.	Master Value on Std. (In °C)	UUC Value (In °C)	Deviation (In °C)	Uncertainty at approx 95% confidence Level and coverage factor k =2 is ± (°C)
1	15.06	15.1	0.04	0.47
2	25.13	25.2	0.07	0.47
3	40.17	40.4	0.23	0.47

Calibration Results (Humidity at 25 °C)

SI. No.	Master Value on Std. (In % RH)	UUC Value (In % RH)	Deviation (In % RH)	Uncertainty at approx 95% confidence Level and coverage factor k =2 is ± (%RH)
1	30.05	30	-0.05	1.77
2	60.09	60	-0.09	1.77
3	90.12	89	-1.12	1.77

Note

- The calibration results reported in this calibration certificate are valid at the time of & under stated condition
- This certificate cannot be reproduced except in full without our prior permission in writing
- This certificate refers only to the particular item(s) submitted for calibration
- UUC - Unit under calibration

----- End of Certificate -----

Bhuma
Calibrated by



Approved by
Anand Chaurasia
Anand Chaurasia
(CEO)



The Accurate Calibration & Medical Solution

CALIBRATION CERTIFICATE

CERTIFICATE OF PERFORMANCE FOR MICROSCOPE

ISSUED TO: M/S KRSNAA DIAGNOSTICS LTD.
ADDRESS: SDII, CIVIL HOSPITAL ROAD SHASTRI NAGAR BATALA (P.B) - 143505
CERTIFICATE NO.: TAC/09-23/KDL/MS-1002
LOCATION/DEPT.: CLINICAL PATHOLOGY
ID NO./BAR CODE: KDSDPBCPMS06
MAKE: LABOMED

DATE: 18.09.2023
MODEL NO.: LX-300
SERIAL NO.: NA

CALIBRATION STANDARD USED

Sl.No.	Master Equipment	Manufacturer	Model No.	Serial No.	Cal. Date	Cal Due Date
1	Electrical Safety Analyzer	DATREND Systems Inc.	vPcd - Mini	VPM18100028	15.10.2022	15.10.2023

ENVIRONMENTAL CONDITIONS

Temperature : $22 \pm 2^{\circ}\text{C}$
Humidity : $50 \pm 10\%RH$

VISUAL TEST

Sr.No.	Tests	Remarks
1	Power Cords, Cable Checking	OK
2	Mains Socket Checking	OK
3	Equipment Type (B,BF,CF)	CF
4	Equipment Class (I, II, III)	I

FUNCTIONAL TEST

Sr.No.	Parameters	Measured Value	Remarks
1	Voltage Between Live and Neutral (VLN)	230.2 V (220-240 V)	OK
2	Voltage Between Live and Earth (VLE)	228.3 V (220-240V)	OK
3	Voltage Between Neutral and Earth (VNE)	0.4 V (0-5V)	OK
4	Load Current (Ia)	0.5 A	OK
5	Enclosure Current	0.6 μA (0-10 μA)	OK
6	Leaking Current (IL)	121 μA (0-300 μA)	OK
7	Time for Testing (t)	15 (min)	OK

Electrical Safety Test Passed as per IEC Standard 60601.1

Calibration Passed YES NO

Calibration Date: 18. 09. 2023

Calibration Due Date: 17. 09. 2024

Remarks:

1. Certificates shall not be reproduced except in full, without the written approval of The Accurate Calibration & Medical Solution.
2. The Calibration result reported in this certificate are valid at the time of Calibration and the stated condition of measurement.

Calibrated by



Approved by

Plot No. 72/G1, Ganesh Vihar Sirsi Road, Jaipur (Raj.) -302012

Mob.: 9672294758

E-mail : theaccuratecalibration@gmail.com



The Accurate Calibration & Medical Solution

CALIBRATION CERTIFICATE

CERTIFICATE OF PERFORMANCE FOR MICROSCOPE

ISSUED TO: M/S KRSNAA DIAGNOSTICS LTD.
ADDRESS: SDH, CIVIL HOSPITAL ROAD SHASTRI NAGAR BATALA (P.B) - 143505
CERTIFICATE NO.: TAC/09-23/KDL/MS-1001
LOCATION/DEPT.: HAEMATOLOGY
ID NO./BAR CODE: KDSDPBHMS05
MAKE: LABOMED

DATE: 18.09.2023
MODEL NO.: LX-300
SERIAL NO.: 21059394

CALIBRATION STANDARD USED

Sl.No.	Master Equipment	Manufacturer	Model No.	Serial No.	Cal. Date	Cal Due Date
1.	Electrical Safety Analyzer	DATREND Systems Inc.	vPed - Mini	VPM18100028	15.10.2022	15.10.2023

ENVIRONMENTAL CONDITIONS

Temperature : $22 \pm 2^{\circ}\text{C}$
Humidity : $50 \pm 10\%RH$

VISUAL TEST

Sr.No.	Tests	Remarks
1	Power Cords,Cable Checking	
2	Mains Socket Checking	OK
3	Equipment Type (B,BF,CF)	OK
4	Equipment Class (I, II, III)	CF
		I

FUNCTIONAL TEST

Sr.No.	Parameters	Measured Value	Remarks
1	Voltage Between Live and Neutral (VLN)	231.6 V (220-240 V)	
2	Voltage Between Live and Earth (VLE)	229.5 V (220-240V)	OK
3	Voltage Between Neutral and Earth (VNE)	0.3 V (0-5V)	OK
4	Load Current (Ia)	0.6 A	OK
5	Enclosure Current	0.8 μA (0-10 μA)	OK
6	Leaking Current (IL)	102 μA (0-300 μA)	OK
7	Time for Testing (t)	15 (min)	OK

Electrical Safety Test Passed as per IEC Standard 60601.1

Calibration Passed YES NO

Calibration Date: 18. 09. 2023

Calibration Due Date: 17. 09. 2024

Remarks:

- 1.Certificates shall not be reproduced except in full,without the written approval of The Accurate Calibration & Medical Solution.
- 2.The Calibration result reported in this certificate are valid at the time of Calibration and the stated condition of measurement.

Calibrated by



Approved by

Plot No. 72/G1, Ganesh Vihar Sirsi Road, Jaipur (Raj.) -302012

Mob.: 9672294758

E-mail : theaccuratecalibration@gmail.com