

A A CALIBRATION PVT. LTD.CALIBRATION CERTIFICATE



CC - 2646

Certificate No	AACPL / 06969F	Field	Elec.Tech.	Group	Page N	lo. 1 of 1
SRF No & Date	230918.1 DT - 18 / 09 / 2023		ULR No.	CC2646230000 06969F		
Calibration Method :	Direct Method		Accuracy			
Company Name	M/s Krsnaa Diagnostics Ltd.		Calibration	Date	18.09.2023	
Address	SDH, Civil Hospital Road,		Recommended	Date for Next Calib.	17.09.2024	
	Shastri Nagar, Batala,		Certificate I	ssue Dt.	20.09.2023	
	Punjab		UUC Condi	tion	Good	

Calibration Instrument Detail

SI	Instrument Name	Range	Least Count	Make / Sl.No.	Party ID / Location
1	Dig. Temp. Controller	2 to 8 °C	1 °C	Samsung	KDSDPBSTRF15
			•	RT28A3022GS/NL/21	Store Room (Refrigerator)

Standard Equipments Used (Traceable to National Standard)

_	Transacta Equipmento Occa / Traccasic to Tracional Standard /					
S	l Instrun	nent Name	Make/SI No	Calibrated By	Cal Certificate No	Cal. Valid Upto
1	Universa	al Calibrator	Radix / 112004873	AACPL	AACPL/00047F	03/04/2024

Reference Standard	Calibration Purpose	Humidity	Temperature	Calibration Performed At
IS : CG-11	Temp. Measurement	(50 ± 20)%RH	(25 ± 10)°C	Site

Calibration Procedure: As per Calibration Procedure Manual AACPL / CP / 03, Section No : CPET - 01

Set Points :- 2, 5 & 8 °C

	Calibration Results					
SI.	SI. Std. Reading on Master UUC Average Reading Deviation Uncertainty at approx 95% c					
No.	in °C	in °C	in °C	Level and coverage factor $k = 2$ is \pm		
1	2.1	2	-0.1	1.62 °C		
2	5.1	5	-0.1	1.62 °C		
3	8.1	9	0.9	1.62 °C		

Note

- 1 The calibration results reported in this calibration certificate are valid at the time of & under stated condition
- 2 This certificate cannot be reproduced except in full without our prior permission in writing
- 3 This certificate refers only to the particular item(s) submitted for calibration
- 4 UUC Unit under calibration

----- End of Certificate -----

Calibrated by



Approyed by

(Tech Manager)

LAB :- Plot No. C-18, Sector-13, First Floor, "Vasundhara, Haat" Vasundhara, Ghaziabad - 201012 Ph.: 0120-4135529, Mob.: 9212172564, 9212172597

E-mail: aacplinfo2005@gmail.com, accplinfo@rediffmail.com