



A A CALIBRATION PVT. LTD.

CALIBRATION CERTIFICATE



Page No. 1 of 1

Certificate No	AACPL / 07062F	Field	Elec Tech	Group		Page No.	1 of 1
SRF No & Date	230926 5	DT - 26 / 09 / 2023		ULR No.	CC2646230000 07062F		
Calibration Method :	Direct Method			Accuracy			
Company Name	M/s Krsnaa Diagnostics Ltd			Calibration Date	26 09 2023		
Address	Mathura Daas Civil Hospital (DH), Main Bazar Moga (PB)			Recommended Date for Next Calib.	25 09 2024		
				Certificate Issue Dt.	25 09 2023		
				UUC Condition	Good		

Calibration Instrument Detail

Instrument Name	Range	Least Count	Make / SI. No	Party ID / Location
1 Dig. Timer (Centrifuge M/c.)	0 - 99 minute	1 minute	Remi	KDDHPBCPCF08 Clinical Pathology

Standard Equipments Used (Traceable to National Standard)

SI	Instrument Name	Make/SI No	Calibrated By	Cal Certificate No	Cal. Valid Upto
1	Dig Time Interval Calibrator	Voltagram /	VTC	CC293522000002078F	27/10/2023

Reference Standard	Calibration Purpose	Humidity	Temperature	Calibration Performed At
IS : 1248	Time Measurement	(50 ± 20) % RH	(25 ± 10) C	Site

Calibration Procedure : As per Calibration Procedure Manual AACPL / CP / 03 Section No CPET - 02

Calibration Result

SI	Selected Range (In s)	Average Read. (Std.) (In s)	Average Read. (UUC) (In s)	Deviation (In s)	Uncertainty at approx 95% confidence Level and coverage factor k = 2 is ±
1	900	900 456	900	-0 456	34 64 s
2	1800	1800 591	1800	-0 591	34 66 s
3	3600	3600 987	3600	-0 987	34 85 s

- Note
- The calibration results reported in this calibration certificate are valid at the time of & under stated condition
 - This certificate cannot be reproduced except in full without our prior permission in writing
 - This certificate refers only to the particular item(s) submitted for calibration
 - UUC - Unit under calibration

----- End of Certificate -----

Approved by