



AACPL CALIBRATION PVT. LTD.

CALIBRATION CERTIFICATE



CC - 2546

Certificate No	AACPL / 07063F	Field	Mech	Group	Speed	Page No.	1 of 1
SRF No & Date	230926 5	DT - 26 / 09 / 2023	ULR No.		CC2546230000 07063F		
Calibration Method :	Comparison Method			Accuracy			
Company Name	M/s Krsnaa Diagnostics Ltd			Calibration Date		26 09 2023	
Address	Mathura Daas Civil Hospital (DH), Main Bazar, Moga (PB)			Recommended Date for Next Calib.		25 09 2024	
				Certificate Issue Dt.		29 09 2023	
				UUC Condition		Good	

Calibration Instrument Detail

Instrument Name	Range	Least Count	Make / Sl. No	Party ID / Location
1 Centrifuge M/c.	0 to 6000 RPM	10 RPM	Remi	KDDHPBCPCF07 Clinical Pathology

Standard Equipments Used (Traceable to National Standard)

Sl	Instrument Name	Make/Sl No	Calibrated By	Cal Certificate No	Cal. Valid Upto
1	Dig. Tachometer	Line Seiki /	AACPL	AACPL / 06363F	31/08/2024

Reference Standard	Calibration Purpose	Humidity	Temperature	Calibration Performed At
IS : 12508 - 2018	RPM Measurement	(50 ± 20)%RH	(25 ± 10) °C	Site

Calibration Procedure : As per Calibration Procedure Manual AACPL / CP / 01B, Section No CP - 08

Calibration Result

Sl	Calib. Gauge Value	Master Gauge Value	Deviation	Uncertainty at approx 95% confidence Level and coverage factor k = 2 is ± (in RPM)
	UUC (in RPM)	(in RPM)	(in RPM)	
1	100	100.1	-0.1	5.88
2	500	500.2	-0.2	5.88
	1000	1000.2	-0.2	5.88
4	3000	3000.4	-0.4	6.15
5	6000	6000.6	-0.6	6.17

Note

- 1 The calibration results reported in this calibration certificate are valid at the time of & under stated condition
- 2 This certificate cannot be reproduced except in full without our prior permission in writing
- 3 This certificate refers only to the particular item(s) submitted for calibration
- 4 UUC - Unit under calibration

----- End of Certificate -----

Signature



Approved by *Signature*

