

ANALYZER CALIBRATION CERTIFICATE

CALIBRATION PROTOCOL

The purpose of this calibration protocol is to define the qualifications and the acceptance standard in order to verify the normal operation and function of the AUTOChem Xpert auto chemistry analyzer in the laboratory. Trained knowledgeable personnel from GENWORKS Health Pvt. Ltd. along with the department personnel will perform and review analyzer calibration protocol as mentioned by the manufacturer. The satisfactorily outcome of this procedures assures that the system functions according to the parameters.

EQUIPMENT INFORMATION:-

Instrument Name : AUTO CHEMISTRY ANALYZER

Model/Type : AUTOChem XPert

Serial No : AES2K2KH04116YA

Installation Date : 13/09/2022

Calibration Done On : 15/09/2023

Next Calibration Due On : 14/09/2024

Laboratory/Hospital Name: MEDRAY CLINICS PVT.LTD

No.962,12th Main Road, Opp.Lakme Salon

Doopanahalli, Indiranagr, Bangalore,

Karnataka - 560008.

Supported By : Genworks Health Pvt Ltd,

Survey No 525/1 & 538/1,

V.S Mani Nagar, Madhavaram,

Chennai-600060



CALIBRATION AND MAINTAINANCE PROCEDURES

MAINTAINANCE:

- Instrument was checked for cleanliness.
- Cuvettes were manually cleaned and performed wash cells(All) operation in maintenance.
- > Wash unit working was checked whether it is dispensing and aspirating Xpert wash and water in the perfect sequence manually.
- Cleaned sample reagent and mixing probes manually with alcohol and then with distilled water also performed probe wash(Detergent) in the maintenance.
- > Performed sample reagent and mixer probes horizontal and vertical movement.
- Air Purge was performed to remove any air bubbles present in the syringes.
- > Instrument mechanism operation check was done and found instrument working in satisfied operation conditions.

CALIBRATION:

- > Input supply and lamp voltage for the machine was found adequate using the multimeter.
- > Cell blank values are measured in the maintenance screen and reports are attached.
- > Cuvette temperature are readings are indicated in the Software
- > Reagent compartments cooling temperature is indicated in LED display behind the machine.
- > Detectors performance were checked by checking the temperature control voltage.
- All parameters Calibration and both level Quality control was done and results found satisfactory



This is to certify that this Analyzer has been inspected and calibrated for following parameters

TEST PARAMETER	TARGET VALUE & RANGE	OBTAINED VALUE
INPUT VOLTAGE	230-240V AC	234V AC
CELL BLANK VAULE	20000- 60000 for all wavelengths	
Cuvette TEMP	37 degree +/- 0.3	37.1 degree
REAGENT COMPARTMENT COOLING CHECK	4 to 14 degrees centigrade	8.2 degree
TEMP CONTROL UNIT	24 +/- 0.2V	
6V LAMP SUPPLY	6 +/- 0.2V	
24V MAIN BOARD SUPPLY	24 +/- 0.3V	

The results are obtained as per specifications & tolerance ranges. Routine chemistry parameters precision study was carried out. The CV's obtained are in acceptable range (< 3.0% CV). Calibrations of routine tests were also done & the results of controls & samples found satisfactory.

Calibration Reports:-

The controls results are obtained as per specifications & tolerance ranges. Verified the calibration & controls results according to the specific limits and considered the satisfactory results herewith certificate that AUTOChem Xpert Analyzer operates correctly according to the instrument specifications.



Calibration details of Standard used

MultimeterMake. : Fluke

Serial No : 40640066WS

Traceability Certificate : HI Tech calibration Services

Certificate # : HT\CC\201107-16\001 &

CR\PCAL\51278Validity: 19-July-2024

Calibration details of Standard used

Temperature SensorMake : PT-100

Instrument ID : WI-CALT-TS02

Traceability Certificate : Northlab (INDIA) PVT.LTD

Certificate # : 0261221-T04 & CC24571400002920F

Validity: 02-DEC-2024

Report Sign Off:

Calibration Done By: Mr.Suresh G J

Designation: Circle Manager

Date & Sign: 15.09.2023

Section Section 201

Note: Supportive date to be attached along with this certificate.

Sample No.:7

Department:

Sample ID:ALANOUD

Bed No.: Age:

Sample Type:Serum

Name:

Gender:Male

Case No.:

Remark:

Doctor: Diagnosis:

Register Date: 22-12-2023 13:53:05

 No.	Test Name	Result	Hint	Unit	Reference range
1 [CI	REA-J1]CRE	0.83		mg/dl	0.40~1.50

Checker:admin Examiner: **BANGALORE** Report Date:22-12-2023 15:08:26

MEDRAY DIAGNOSTICS

INDIRANAGAR

Sample No.:8 Name:

Department:

Remark:

Sample ID:ALANOUD

Doctor:

Gender:Male

Bed No.: Age:

Sample Type:Serum

Case No.:

Diagnosis:

Register Date: 22-12-2023 13:53:05

No.	Test Name	Result	Hint	Unit	Reference range
 1 [CI	REA-J11CRE	1.32		mg/dl	0.40~1.50

Sample No.:9

Sample ID:ALANOUD

Gender:Male

Bed No.: Sample Type:Serum

Age:

Case No.:

Department: Remark:

Name:

Doctor:

Diagnosis:

Register Date:22-12-2023 13:53:05

No.	Test Name	Result	Hint	Unit	Reference range
1	[CREA-J1]CRE	1.38		mg/dl	0.40~1.50

Checker:admin Examiner: **BANGALORE** Report Date:22-12-2023 15:08:26

MEDRAY DIAGNOSTICS

INDIRANAGAR

Sample No.:10

Name:

Remark:

Sample ID:ALANOUD

Gender:Male

Doctor:

Bed No.:

Age:

Sample Type:Serum

Case No.:

Department:

Diagnosis:

Register Date: 22-12-2023 13:53:05

No.	Test Name	Result	Hint	Unit	Reference range
1	[CREA-J1]CRE	0.89		mg/dl	0.40~1.50

Sample No.:11

Department:

Sample ID:ALANOUD

Doctor:

Bed No.: Age:

Sample Type:Serum

Gender:Male Name:

Diagnosis:

Case No.:

Remark:

Register Date:22-12-2023 13:53:05

No.	Test Name	Result	Hint	Unit	Reference range
1 [C	REA-J1]CRE	0.76		mg/dl	0.40~1.50

Checker:admin Examiner: **BANGALORE** Report Date:22-12-2023 15:08:26

MEDRAY DIAGNOSTICS

INDIRANAGAR

Sample No.:12

Sample ID:ALANOUD

Sample Type:Serum

Name:

Gender:Male

Age:

Case No.:

Department: Remark:

Doctor:

Diagnosis:

Bed No.:

Register Date: 22-12-2023 13:53:05

No.	Test Name	Result	Hint	Unit	Reference range
1	[CREA-J1]CRE	0.77		mg/dl	0.40~1.50

Age:

Sample No.:13

Sample ID:ALANOUD

Gender:Male

Sample Type:Serum

Case No.:

Department: Remark:

Name:

Doctor:

Diagnosis:

Bed No.:

Register Date: 22-12-2023 13:53:05

N	lo.	Test Name	Result	Hint	Unit	Reference range
1	[CR	EA-J1]CRE	0.80		mg/dl	0.40~1.50

Checker:admin Examiner: **BANGALORE** Report Date:22-12-2023 15:08:26

MEDRAY DIAGNOSTICS

INDIRANAGAR

Sample No.:14

Name:

Remark:

Sample ID:ALANOUD

Gender:Male

Age:

Bed No.:

Sample Type:Serum

Case No.:

Department: Doctor:

Diagnosis:

Register Date:22-12-2023 13:53:05

No. Test Name Result Hint Unit Reference range [CREA-J1]CRE 0.80 0.40~1.50 1 mg/dl

Sample No.:15

Sample ID:ALANOUD

Gender:Male

Doctor:

Bed No.: Sample Type:Serum

Age:

Case No.:

Department: Remark:

Name:

Diagnosis:

Register Date:22-12-2023 13:53:05

No.	Test Name	Result	Hint	Unit	Reference range
1	[CREA-J1]CRE	0.82		mg/dl	0.40~1.50

Checker:admin Examiner: **BANGALORE** Report Date:22-12-2023 15:08:26

MEDRAY DIAGNOSTICS

INDIRANAGAR

Sample No.:16

Department:

Name:

Sample ID:ALANOUD

Gender:Male

Doctor:

Bed No.: Age:

Sample Type:Serum

Case No.:

Diagnosis:

Remark:	Remark: Register Date:22-12-2023 13:53:0			
No. Test Name	Result H	int Unit	Reference range	
1 [CREA-J1]CRE	0.81	mg/dl	0.40~1.50	

Sample No.:2

Sample ID:MD ABDUL MOMIN

Bed No.: Sample Type:Serum

Gender:Male

Age:

Case No.:

Department: Remark:

Name:

Doctor:

Diagnosis:

Register Date:19-12-2023 12:31:30

No.	Test Name	Result	Hint	Unit	Reference range
1 [0	GLU3]GLUCOSE	109.4		mg/dl	70.0~140.0

Checker:admin Examiner: **BANGALORE** Report Date:19-12-2023 13:14:39

MEDRAY DIAGNOSTICS INDIRANAGAR

Sample No.:3

Sample ID:MD ABDUL MOMIN

Bed No.: Age:

Sample Type:Serum

Name:

Gender:Male

Case No.:

Department:

Remark:

Doctor:

Diagnosis:

No	Test Name	Result	Hint	Unit	Reference range
1	[GLU3]GLUCOSE	101.6		mg/dl	70.0~140.0

Sample No.:4

Sample ID:MD ABDUL MOMIN

Bed No.:

Sample Type:Serum

Name:

Gender:Male

Age:

Case No.:

Department: Remark:

Doctor:

Diagnosis:

Register Date:19-12-2023 12:31:30

No.	Test Name	Result	Hint	Unit	Reference range
1 [0	GLU3]GLUCOSE	101.0		mg/dl	70.0~140.0

Checker:admin Examiner: **BANGALORE** Report Date:19-12-2023 13:14:39

MEDRAY DIAGNOSTICS INDIRANAGAR

Sample No.:5

Sample ID:MD ABDUL MOMIN

Bed No.:

Sample Type:Serum

Name:

Gender:Male

Age:

Case No.:

Department:

Remark:

Doctor:

Diagnosis:

No	Test Name	Result	Hint	Unit	Reference range
1	[GLU3]GLUCOSE	101.1		mg/dl	70.0~140.0

Sample No.:6

Department:

Sample ID:MD ABDUL MOMIN

Gender:Male Name:

Doctor:

Bed No.: Age:

Sample Type:Serum Case No.:

Remark:

Diagnosis:

Register Date:19-12-2023 12:31:30

No.	Test Name	Result	Hint	Unit	Reference range
1 [0	GLU3]GLUCOSE	98.9		mg/dl	70.0~140.0

Checker:admin Examiner: **BANGALORE** Report Date:19-12-2023 13:14:39

MEDRAY DIAGNOSTICS INDIRANAGAR

Sample No.:7 Name:

Department: Remark:

Sample ID:MD ABDUL MOMIN

Gender:Male

Doctor:

Bed No.:

Age:

Diagnosis:

Sample Type:Serum

Case No.:

No	Test Name	Result	Hint	Unit	Reference range
1	[GLU3]GLUCOSE	99.3		mg/dl	70.0~140.0

Sample No.:8

Sample ID:MD ABDUL MOMIN

Gender:Male

Age:

Sample Type:Serum

Name:

Bed No.:

Department: Remark:

Doctor:

Diagnosis:

Register Date:19-12-2023 12:31:30

Case No.:

No.	Test Name	Result	Hint	Unit	Reference range
1	[GLU3]GLUCOSE	100.0		mg/dl	70.0~140.0

Checker:admin Examiner: **BANGALORE** Report Date:19-12-2023 13:14:39

MEDRAY DIAGNOSTICS INDIRANAGAR

Sample No.:9

Sample ID:MD ABDUL MOMIN

Bed No.:

Sample Type:Serum

Name:

Gender:Male

Age:

Case No.:

Department: Remark:

Doctor:

Diagnosis:

	No.	Test Name	Result	Hint	Unit	Reference range
•	1 [G	LU3]GLUCOSE	98.7		mg/dl	70.0~140.0

Sample No.:10

Sample ID:MD ABDUL MOMIN

Bed No.:

Sample Type:Serum

Name:

Gender:Male

Age:

Department: Remark:

Doctor: Diagnosis:

Register Date:19-12-2023 12:31:30

Case No.:

No.	Test Name	Result	Hint	Unit	Reference range
1 [0	GLU3]GLUCOSE	100.8		mg/dl	70.0~140.0

Checker:admin Examiner: **BANGALORE** Report Date:19-12-2023 13:14:39

MEDRAY DIAGNOSTICS INDIRANAGAR

Sample No.:11

Sample ID:MD ABDUL MOMIN

Age:

Sample Type:Serum

Name:

Gender:Male

Bed No.:

Case No.:

Department: Remark:

Doctor:

Diagnosis:

No	. Test Name	Result	Hint	Unit	Reference range
1	[GLU3]GLUCOSE	100.8		mg/dl	70.0~140.0

Auto Chem Xpert ANALYSER - PRECISION data						
CUSTOMER NAME	Medray Clinics Pvt Ltd					
	Parameters					
S.No	Glucose	Creatinine				
1	109.4	0.83				
2	101.6	1.32				
3	101	1.38				
4	101.1	0.89				
5	98.9	0.76				
6	99.3	0.77				
7	100	0.8				
8	98.7	0.8				
9	100.8	0.82				
10	100.8	0.81				
Mean	101.16	0.92				
SD	3.06	0.23				
CV%	3.03	25.15				