



PROFICIENCY TESTING REPORT
ISHTM-AIIMS EXTERNAL QUALITY ASSURANCE PROGRAMME
 NABL accredited program as per ISO/IEC 17043:2010 standard
 Organized By Department of Hematology, AIIMS, New Delhi-110029



Duration of stability testing - minimum upto 8 days at ambient temp. after dispatch of specimens

EQAP CODE No. : 2405

Distribution No.: 161-E

Month/Year: September/2023

Instrument ID: 13575

Name & Contact No. of PT Co-ordinator: Dr. Manoranjan Mahapatra (Prof. & Head), Hematology, AIIMS, Delhi,
 Tel: 9013085730 , E-Mail : accuracy2000@gmail.com

Date of issue & status of the report: 30-10-2023[Final].

CBC and Retic Assessment

Test Parameters	S.No.	Among Lab (Accuracy Testing)						Within Lab (Precision Testing)			
		Your Result 1	Your Result 2	Your Results Sum of 2 Value	Consensus result sum of 2 values (Assigned Value)	Uncertainty of Assigned Values	Z Score	Yours Results Diff. of 2 Values	Consensus Result Diff. of 2 values (Assigned Value)	Uncertainty of Assigned Values	Z Score
WBC x10 ³ /µl	1	3.83	3.76	7.59	6.88	0.038	0.74	0.07	0.1	0.005	-0.40
RBC x10 ⁶ /µl	1	4.37	4.32	8.69	8.82	0.011	-0.50	0.05	0.04	0.002	0.27
Hb g/dl	1	12.8	12.6	25.4	25.75	0.029	-0.47	0.2	0.1	0.008	0.67
HCT%	1	43.7	43	86.7	81.6	0.186	1.06	0.7	0.4	0.025	0.81
MCV-fl	1	100	99.5	199.5	184.6	0.302	1.69	0.5	0.3	0.024	0.45
MCH-Pg	1	29.3	29.2	58.5	58.2	0.070	0.18	0.1	0.2	0.016	-0.34
MCHC-g/dl	1	29.3	29.3	58.6	62.8	0.146	-1.14	0	0.3	0.022	-1.01
Plt. x10 ³ /µl	1	124	109	233	306	1.568	-1.81	15	5	0.336	1.69
Retic %	2										

P.S . Assesment

YOUR REPORT		CONSENSUS REPORT
DLC%	3	Nrbcs= , Poly=78 L=2, E=, Mono/Promono=8 , B1=2 P.M.=, Mye=9, Meta=, Other=
RBC Morphology	3	Poly: 46 - 62, Myelo: 11 - 20, Meta: 7- 15, Promyelo: 2-7, Lympho: 2- 5, Blast: 2-4, Eosino: 1-3, nRBC/Mono, Baso: 0-6
Diagnosis	3	hyperleukocytosis with marked myeloid left shift andbarophilic
		Predominantly: Normocytic/Normochromic; Moderate: Anisocytosis, Hypochromic, Mild: Poikilocytosis
		chronic myeloid leukemia chronic phase
		Chronic Myeloid Leukemia (Chronic Phase)

COMBINED DATA VALUES OF TOTAL PARTICIPANTS

Test parameters	S.No.	Total participants covered in the current dist. 161--E	Total No. responded	% of Labs with Z Score 0-2		% of Labs with Z Score 2-3		% of Labs with Z Score >3	
				Among labs	Within lab	Among labs	Within lab	Among labs	Within lab
WBC x10³/µl	1	314	314	88.54	92.99	5.41	3.82	6.05	3.19
RBC x10⁶/µl	1	314	314	85.67	90.45	7.01	2.55	7.32	7
Hb g/dl	1	314	314	86.31	90.45	4.46	5.1	9.23	4.45
HCT%	1	314	313	87.86	92.33	7.67	2.56	4.47	5.11
MCV-fl	1	314	314	92.99	93.63	2.55	3.5	4.46	2.87
MCH-Pg	1	314	314	85.35	92.04	6.69	4.46	7.96	3.5
MCHC-g/dl	1	314	314	88.85	86.31	7.01	2.87	4.14	10.82
Plt. x10³/µl	1	314	314	90.76	91.72	6.37	4.46	2.87	3.82
ReticCount%	2	314	263	92.4	80.99	4.18	0.76	3.42	18.25
PS Assessment	3	314	281	Satisfactory :94.92%, Borderline Sat. :2.54%, Unsatisfactory :2.54%					

***Comments:**

1). **Among Lab (EQA) : Results acceptable.**

2). **Within Lab (IQA) : Precision acceptable.**

Note-1: EQA (External Quality Assurance) : Your Performance among various of participating labs in PT, to determine the accuracy of your results.

IQA (Internal Quality Assurance) : Your Performance of comparison of two consecutive measurement values within your lab to test the precision of your autoanalyzer.

Note-2: Z score among & within lab were calculated, as per to ISO/IEC 13528:2015 standard. Z score among lab (EQA)= (Your Result Sum of two values - Consensus Result sum of two values)/(Normalised IQR)

Z score within lab (IQA)= (Your Result Difference of two values - Consensus Result difference of two values)/(Normalised IQR)

IQR = Quartile 3 - Quartile 1 of participant data, Normalised IQR = 0.7413 x IQR

Note-3: Z score 0 to ± 2 : Acceptable, Z score ± 2 to ± 3 :Warning Signal, Z score $> \pm 3$: Unacceptable [As per ISO/IEC 13528:2015 standard]

Note-4: Z score value between "0 to ± 2 " are texted in green colour. Z score value between " ± 2 to ± 3 " are texted in orange colour. Z score value $> \pm 3$ are texted in red colour.

Note-5: Homogeneity and stability testing of PT sample were done as per ISO 13528:2015 standard. To pass homogeneity test, between sample SD (Ss) should be smaller than the check value (0.3*SDPA). To pass the stability test, average difference in measurement values of first and last day sample ($\bar{x}-\bar{y}$) should be smaller than the check value (0.3*SDPA).

Note-6: ISHTM-AIIMS-EQAP does not subcontract any task of its scheme

Note-7: Participants are free to use methods/analyzer of their own choice.

Note-8: Proficiency testing (PT) samples are sent quarterly to each participant.

Note-9: All the necessary details regarding design and implementation of PT, are provided in the instruction sheet as well as on programme's website www.ishtmaimseqap.com.

Note 10: Reports are kept confidential.

Report authorized by,



Dr. Manoranjan Mahapatra (Prof. & Head)

PT Co-ordinator: ISHTM-AIIMS-EQAP

Department of Hematology, AIIMS, New Delhi

-----End Of Report-----