

HORIBA
Medical

HORIBA India Private Limited

246, Okhla Industrial Estate Phase-III,
New Delhi 110020, India
Tel : +91 (11) 4646 5000 / 4669 5001
Fax : +91 (11) 4646 5020
http : //www.horiba.com
CIN : U73100DL2006PTC153232

HIN/MED/2023-2024 /10556540
30TH OCT 2023

CALIBRATION CERTIFICATE

This is to certify that the Hematology Analyzer Horiba YUMIZEN H550 bearing serial number 109YAXH03509 installed at **KRSNAA DIAGNOSTIC LAB BATALA** as calibrated on 30TH OCT 2023.

Calibrator: ABX MINOCAL

Lot NO : CX487

Expiry Date : 05th NOV 2023

The reports of Blank Cycle, Repeatability and Calibration Values were all found in acceptable range.

Next calibration cycle is due on 29TH OCT 2024



Shrish Dixit

(Head- Products &
Marketing) For **Horiba
India Pvt. Ltd.**

246, Okhla Industrial
Estate Phase-III, New Delhi
110020, India
Tel : +91 (11) 4546 5000 / 46695001
Fax : +91 (11) 4646 5020
http : //www.horiba.com
CIN : U73100DL2006PTC1M23