



## CERTIFICATE OF CALIBRATION

**Customer Name** : Redcliffe Diagnostic Lab, Jalandhar

**Model** : Laura

**Serial No.** : 1065-1-202209

**Calibration Date** : 11/05/2023

**Calibration Due Date** : 10/05/2024

This is to certify that we have calibrated the Urine Chemistry Analyzer, **Model: Laura** bearing Sr. No. 1065-1-202209 and on inspection of the instrument it is observed that the results are well within the range and instrument is working satisfactorily.

Thanking you,

For Transasia Bio-Medicals Ltd.



**Rashpal Singh Rana**  
**Sr. Regional Service Manager**

M: +91-9316004670

Tel Fax No: +91-172-4643837

E mail: [r.rana@transasia.co.in](mailto:r.rana@transasia.co.in)

DEKAPHAN LAURA

Seq.No: 0001


BIORAD N

28.12.2023

08:53 AM



BLD	NEG
BIL	NEG
UBG	NORM
KET	NEG
GLU	NEG
PRO	NEG
pH	6
NIT	NEG
LEU	NEG
SG	1.025



QC TEST: OK

LOT: 2308055 EXP: 31.08.2024

<Admin>

DEKAPHAN LAURA

Seq.No: 0004

BIORAD P

28.12.2023

09:13 AM

.....

*BLD	10	Ery/u1
*BIL	3	mg/dl
*UBG	1	mg/dl
*KET	52	mg/dl
*GLU	1000	mg/dl
*PRO	100	mg/dl
PH	9	
*NIT	POS	
*LEU	75	Leu/u1
SG	1.000	

-----

QC TEST: OK

: 2308055 EXP: 31.08.2024

<Admin>

*(Handwritten signature and blue circular stamp)*

