



TBS INDIA

TELEMATIC AND BIOMEDICAL SERVICES PVT. LTD.



CERTIFICATE OF PERFORMANCE FOR REFRIGERATOR

ISSUED TO : «MODERN DIAGNOSTIC CENTER»
CERTIFICATE NO : «TBS/CAL/MDC/2023/22»
LOCATION/DEPT. : «MAIN LABORATORY» DATE : «04/12/23»
ASSET DESCRIP. : «REFRIGERATOR» MODEL NO : «VGID295»
MANUFACTURER : «FRIGOGLASS INDIA » SERIAL NO : «IN6321260018»

CALIBRATION STANDARD USED

S.No	Asset No	Test Equipment	Manufacturer	Model No	Serial No	Cal. Date	Cal. Due Date
1	TBS/TE/0281	DIGITAL THERMOMETER	WACO	201	130230	19.04.23	19.04.24
2	TBS/TE/0185	ELECTRICAL SAFETY ANALYZER	RIGEL MEDICAL	288+	25J-0310	26.04.23	25.04.24

TEST REPORT

S.No	Parameters	Unit	Set Value	Measured Value	Acceptable Range	Remarks
1	TEMPERATURE	C	2	«2»	± 0.3 °C	PASS
			4	«3.7»	± 0.3 °C	PASS
			6	«5.9»	± 0.3 °C	PASS
			NIL	NIL	NIL	NIL

Electrical Safety Test Passed as per IEC Standard 60601.1

Calibration Passed: YES NO

Calibration Done Date: «01.12.23»

Calibration Due Date: «01.12.24»

Calibrated By: «SHAHID HUSSAIN

Approved By: Dr Mubashir Hassan



The above mentioned instrument has been certified with accuracy guidelines traceable to the National / International Standards
5th Floor, Arden Fair, Opp Benniganahalli Ring Road Flyover, Pai Layout, Near Tin Factory, Old Madras Road,
Bangalore – 560016, Tel: +91 80 40545050 Fax: +91 80 40545055
Email: info@tbs-india.com Website: www.tbs-india.com