



TBS INDIA

TELEMATIC AND BIOMEDICAL SERVICES PVT. LTD.

CERTIFICATE OF PERFORMANCE FOR CENTRIFUGE

ISSUED TO : MODERN DIAGNOSTIC CENTER.
CERTIFICATE NO : «TBS/CAL/MDC/2023/21»
LOCATION/DEPT. : «MAIN LABORATORY» DATE : «04/12/2023»
ASSET DESCRIP. : «CENTRIFUGE 4K RPM» MODEL NO : «R8C»
MANUFACTURER : «REMI» SERIAL NO : «2EHN28426»

CALIBRATION STANDARD USED

S.No	Asset No	Test Equipment	Manufacturer	Model No	Serial No	Cal. Date	Cal. Due Date
1	«TBS/TE/0385»	«TACHOMETER»	«METRAVI»	«NCTM-2000»	171103535	«19.4.23»	«19.4.24»
2	«TBS/TE/0185»	«EST ANALYZER»	«RIGEL»	«288+»	«25J-0310»	«26.4.23»	«25.4.24»

TEST REPORT

S.No	Parameters	Unit	Set Value	Measured Value	Acceptable Range	Remarks
1	SPEED	RPM	1000	«997»	± 5%	PASS
			2000	2005	± 5%	PASS
			3000	3000	± 5%	PASS
			4000	4025	± 5%	PASS
			NIL	NIL	NIL	NIL

Electrical Safety Test Passed as per IEC Standard 60601.1

Calibration Passed: YES NO

Calibration Done Date: «01.12.2023»

Calibration Due Date: «01.12.2024»

Calibrated By: «Shahid Hussain»

Approved By: _Dr Mubashir Hassan



The above mentioned instrument has been certified with accuracy guidelines traceable to the National / International Standards
5th Floor, Arden Fair, Opp Benniganahalli Ring Road Flyover, Pai Layout, Near Tin Factory, Old Madras Road,
Bangalore – 560016, Tel: +91 80 40545050 Fax: +91 80 40545055
Email: info@tbs-india.com Website: www.tbs-india.com