

ULR NO: CF48JKDG8949 ISSUE DATE: 25.12.2023	CALIBRATION DONE ON : 25.12.2023 CALIBRATION VALID UPTO : 04.12.2024	SRF. & DATE: 3050/935 & 25.12.2023 PAGE : 1 OF 1
CALIBRATED FOR M/S. VEDANTA HOSPITAL AZAMGARH, UTTAR PRADESH		

IDENTIFICATION ON UUC (UNIT UNDER CALIBRATION)

INSTRUMENT NAME	MICRO PIPETTE	SIZE/RANGE	100 -1000 µl
LEAST COUNT	10 µl	CONDITION OF ITEM	GOOD
MANUFACTURE BY	DRAGON LAB	CALIBRATION PERFORMED AT	SITE
SR NO	KW02551	LOCATION	LAB
CALIBRATION ID	ID/LAB/PP-01	REF. STANDARD	ISO-8655-6
ENVIRONMENT CONDITION: TEMPERATURE: 25 ± 5, RELATIVE HUMIDITY: 55 ± 15% RH			

STANDARD AND MAJOR EQUIPMENTS

STANDARD & EQUIPMENT	RANGE	CALIBRATING AGENCY	TRACEBILITY CERT. NO.	CERTIFICATION DATE	VALID UPTO
STANDARD WEIGHT	1 MG to 200 g	Weightronics	SDJGH48937/22	15/11/2023	14/11/2024

CALIBRATION RESULT AT 27 °C

NORMAL VALUE UUC (µl)	OBSERVED VALUE (µl)	ERROR (µl)	RELATED UNCERTAINTY ON C.F.(K=2) AT 95% CONFIDENCE LEVEL ±
100	100.49	-0.40	0.38 µl
200	200.56	-0.56	0.38 µl
500	501.83	-0.83	0.38 µl
1000	1003.91	-3.91	0.38 µl

.....END OF REPORT.....



Calibration Engineer's Signature