



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME
MONTHLY SUMMARY REPORT - JUNE 2023



PC-1024

Lab Name **DELVIN PATHOLOGY LABORATORY** Lab No **18304**
 Constituent Group **Chemistry II** Date of Result Entered : **15/06/2023**
 PT item **Lyophilized human serum based** Date of Report Published : **04/07/2023**

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE II	GOD-POD II	Transasia / Erba	978	272.81	10.91	29.76	276 mg/dL	0.11	1.90
2	UREA II	Urease UV / GLDH II	Transasia / Erba	750	47.46	9.63	4.57	50 mg/dL	0.56	0.33
3	CREATININE II	Jaffes Kinetic-Alkaline Picrate II	Transasia / Erba	603	1.08	13.93	0.15	0.8 mg/dL	-1.85	0.01
4	T.BILIRUBIN II	Diazonium Salt (Colorimetric) / Jendrassik II	Transasia / Erba	903	4.75	11.67	0.56	5.4 mg/dL	1.17	0.04
5	T-PROTEIN II	Biuret - Colorimetric II	Transasia / Erba	845	6.08	8.36	0.51	6.5 g/dL	0.83	0.03
6	ALBUMIN II	BCG - Colorimetric II	Transasia / Erba	843	3.54	7.91	0.28	3.6 g/dL	0.21	0.02
7	URIC ACID II	Enzymatic / Uricase Colorimetric II	Transasia / Erba	839	4.47	15.69	0.70	3.7 mg/dL	-1.10	0.05
8	CHOLESTROL II	CHOD-PAP II	Transasia / Erba	1000	124.05	12.14	15.06	130 mg/dL	0.40	0.95
9	TRIGLYCERIDE II	GPO-PAP / Enzymatic Colorimetric / End Point II	Transasia / Erba	918	145.60	15.05	21.92	190 mg/dL	2.03	1.45

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ± 1.01 to ± 2.00	Good.
Within ± 2.01 to ± 2.99	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :

DELVIN PATHOLOGY LABORATORY
 Shop No FF01 Second Floor, Parsvnath Plaza, G Block, Pocket G, Sector 27
 NOIDA
 UTTAR PRADESH 201301

Coordinator Contact Details:
 Email: clinqc@cmcvellore.ac.in
 Contact Number: 0416-2283102

Pamela Christudoss
 Dr. Pamela Christudoss
 CMC EQAS Coordinator
 Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

CMC EQAS does not sub contract any components

***** End of Report *****