



# CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY  
CMC EXTERNAL QUALITY ASSURANCE SCHEME  
MONTHLY SUMMARY REPORT - OCTOBER 2023



PC-1024

Lab Name **DELVIN PATHOLOGY LABORATORY** Lab No **18304**  
 Constituent Group **Chemistry II** Date of Result Entered : **12/10/2023**  
 PT item **Lyophilized human serum based** Date of Report Published : **04/11/2023**

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE II	GOD-POD II	Transasia / Erba	1091	128.85	9.70	12.49	122 mg/dL	-0.55	0.76
2	UREA II	Urease UV / GLDH II	Transasia / Erba	806	88.19	10.40	9.17	73.9 mg/dL	-1.56	0.65
3	CREATININE II	Jaffes Kinetic-Alkaline Picrate II	Transasia / Erba	659	7.31	12.63	0.92	7.8 mg/dL	0.53	0.07
4	T.BILIRUBIN II	Diazonium Salt (Colorimetric) / Jendrassik II	Transasia / Erba	918	8.42	11.16	0.94	9.04 mg/dL	0.66	0.06
5	T-PROTEIN II	Biuret - Colorimetric II	Transasia / Erba	894	5.52	8.77	0.48	5.5 g/dL	-0.04	0.03
6	ALBUMIN II	BCG - Colorimetric II	Transasia / Erba	898	3.28	8.74	0.29	3.4 g/dL	0.42	0.02
7	URIC ACID II	Enzymatic / Uricase Colorimetric II	Transasia / Erba	891	4.99	12.10	0.60	5.2 mg/dL	0.35	0.04
8	CHOLESTROL II	CHOD-PAP II	Transasia / Erba	909	103.00	13.09	13.48	115 mg/dL	0.89	0.89
9	TRIGLYCERIDE II	GPO-PAP / Enzymatic Colorimetric / End Point II	Transasia / Erba	865	89.20	16.03	14.30	60 mg/dL	-2.04	0.97

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within $\pm 1.01$ to $\pm 2.00$	Good.
Within $\pm 2.01$ to $\pm 2.99$	Accept with caution. Warning Signal.
Beyond $\pm 3.0$	Unacceptable performance. Action Signal.

## LAB ADDRESS :

**DELVIN PATHOLOGY LABORATORY**  
 Shop No FF01 Second Floor, Parsvnath Plaza, G Block, Pocket G, Sector 27  
 NOIDA  
 UTTAR PRADESH 201301

Coordinator Contact Details:  
 Email: [clinqc@cmcvellore.ac.in](mailto:clinqc@cmcvellore.ac.in)  
 Contact Number: 0416-2283102

*Pamela Christudoss*  
 Dr. Pamela Christudoss  
 CMC EQAS Coordinator  
 Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.  
 Data in CMC EQAS reports is confidential  
 CMC EQAS does not sub contract any components

\*\*\*\*\* End of Report \*\*\*\*\*