



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME
MONTHLY SUMMARY REPORT - NOVEMBER 2023



PC-1024

Lab Name **DELVIN PATHOLOGY LABORATORY** Lab No **18304**
Constituent Group **Chemistry II** Date of Result Entered : **12/11/2023**
PT item **Lyophilized human serum based** Date of Report Published : **05/12/2023**

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE II	GOD-POD II	Transasia / Erba	1079	244.92	10.42	25.53	241 mg/dL	-0.15	1.55
2	UREA II	Urease UV / GLDH II	Transasia / Erba	808	68.11	9.41	6.41	68 mg/dL	-0.02	0.45
3	CREATININE II	Jaffes Kinetic-Alkaline Picrate II	Transasia / Erba	661	4.57	11.48	0.52	4.5 mg/dL	-0.13	0.04
4	T.BILIRUBIN II	Diazonium Salt (Colorimetric) / Jendrassik II	Transasia / Erba	959	1.11	24.30	0.27	1.4 mg/dL	1.08	0.02
5	T-PROTEIN II	Biuret - Colorimetric II	Transasia / Erba	913	4.78	11.66	0.56	4.6 g/dL	-0.32	0.04
6	ALBUMIN II	BCG - Colorimetric II	Transasia / Erba	902	2.87	9.16	0.26	2.9 g/dL	0.11	0.02
7	URIC ACID II	Enzymatic / Uricase Colorimetric II	Transasia / Erba	944	8.28	12.24	1.01	7.5 mg/dL	-0.77	0.07
8	CHOLESTROL II	CHOD-PAP II	Transasia / Erba	985	93.91	12.00	11.27	80 mg/dL	-1.23	0.72
9	TRIGLYCERIDE II	GPO-PAP / Enzymatic Colorimetric / End Point II	Transasia / Erba	855	76.30	22.68	17.30	79 mg/dL	0.16	1.18

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ± 1.01 to ± 2.00	Good.
Within ± 2.01 to ± 2.99	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :

DELVIN PATHOLOGY LABORATORY
Shop No FF01 Second Floor, Parsvnath Plaza, G Block, Pocket G, Sector 27
NOIDA
UTTAR PRADESH 201301

Coordinator Contact Details:
Email: clinqc@cmcvellore.ac.in
Contact Number: 0416-2283102

Pamela Christudoss
Dr. Pamela Christudoss
CMC EQAS Coordinator
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
Data in CMC EQAS reports is confidential
CMC EQAS does not sub contract any components

***** End of Report *****