



AACPL CALIBRATION PVT. LTD.

CALIBRATION CERTIFICATE

7.8F-01



CC - 2646

Certificate No	AACPL / 07108F	Field	Thermal	Group	Temp.	Page No.	1 of 1
SRF No & Date	230929.7 DT - 29 / 09 / 2023		ULR No.	CC2646230000 07108F			
Calibration Method :	Comparison Method		Accuracy				
Company Name	M/s Krsnaa Diagnostics Ltd.		Calibration Date	29.09.2023			
Address	Shri Prabodh Chander District Hospital Shahpur Road, Pathankot (PB)		Recommended Date for Next Calib.	28.09.2024			
			Certificate Issue Dt.	03.10.2023			
			UUC Condition	Good			

Calibration Instrument Detail

SI	Instrument Name	Range	Least Count	Make/SI.No.	Party ID / Location
1	Incubator	0 to 110 °C	1 °C	Arba Healthcare ----	KDDHPBCCIN13 Clinical Chemistry

Standard Equipments Used (Traceable to National Standard)

SI	Instrument Name	Make/SI No	Calibrated By	Cal Certificate No	Cal. Valid Upto
1	Digital Multimeter	Gwinstek EL 160577	AACPL	AACPL/04256F	11/07/2024
2	RTD (4 Wire) Pt-100	Tempens / 23000375	Tempens	TL/022/1114.1.1	06/10/2023

Reference Standard	Calibration Purpose	Humidity	Temperature	Calibration Performed At
DKD-R5-7	Temp. Measurement	(50 ± 20)%RH	(25 ± 10)°C	Site

Calibration Procedure

Comparison With Standard Sensors as per Calibration Procedure Manual AACPL / CP / 02, Section No : CPTH - 09

Calibration Points :- 10 , 60 & 110 °C

Calibration Results				
SI No.	Master Value	UUC Value	Deviation	Uncertainty at approx 95% confidence Level and coverage factor k =2 is ± (°C)
	In °C	In °C	In °C	
1	10.09	10	-0.09	1.24
2	60.15	60	-0.15	1.24
3	110.22	110	-0.22	1.24

Note

- 1 The calibration results reported in this calibration certificate are valid at the time of & under stated condition
- 2 This certificate cannot be reproduced except in full without our prior permission in writing
- 3 This certificate refers only to the particular item(s) submitted for calibration
- 4 UUC - Unit under calibration
- 5 Temperature Scale: International Temperature Scale Of 1990 (ITS-90)

----- End of Certificate -----

Awasb
Calibrated by



Approved by
[Signature]
Anand Chaurasia
(CEO)