



A A CALIBRATION PVT. LTD.

CALIBRATION CERTIFICATE

7.8F-01



CC - 2646

Certificate No	AACPL / 07107F	Field	Mech.	Group	Speed	Page No.	1 of 1
SRF No & Date	230929.7 DT - 29 / 09 / 2023	ULR No.		CC2646230000 07107F			
Calibration Method :	Comparison Method			Accuracy			
Company Name	M/s Krsnaa Diagnostics Ltd.	Calibration Date		29.09.2023			
Address	Shri Prabodh Chander District Hospital Shahpur Road, Pathankot (PB)	Recommended Date for Next Calib.		28.09.2024			
		Certificate Issue Dt.		03.10.2023			
		UUC Condition		Good			

Calibration Instrument Detail

SI	Instrument Name	Range	Least Count	Make / SI. No	Party ID / Location
1	Centrifuge M/c.	0 to 3500 RPM	100 RPM	Wisdom -----	KDDHPBCPCF07 Clinical Pathology

Standard Equipments Used (Traceable to National Standard)

SI	Instrument Name	Make/SI No	Calibrated By	Cal Certificate No	Cal. Valid Upto
1	Dig. Tachometer	Line Seiki / -----	AACPL	AACPL / 06363F	31/08/2024

Reference Standard	Calibration Purpose	Humidity	Temperature	Calibration Performed At
IS : 12508 - 2018	RPM Measurement	(50 ± 20)%RH	(25 ± 10) °C	Site

Calibration Procedure : As per Calibration Procedure Manual AACPL / CP / 01B, Section No : CP - 08

Calibration Result

SI	Calib. Gauge Value	Master Gauge Value	Deviation	Uncertainty at approx 95% confidence Level and coverage factor $k = 2$ is \pm (in RPM)
	UUC (in RPM)	(in RPM)	(in RPM)	
1	100	100.6	-0.6	57.75
2	500	501.3	-1.3	57.75
3	1000	1001.8	-1.8	57.75
4	3000	3002.7	-2.7	57.77
5	6000	6003.5	-3.5	57.78

Note

- The calibration results reported in this calibration certificate are valid at the time of & under stated condition
- This certificate cannot be reproduced except in full without our prior permission in writing
- This certificate refers only to the particular item(s) submitted for calibration
- UUC - Unit under calibration

----- End of Certificate -----

Amay
Calibrated by



Approved by
[Signature]
Anand Chaurasia
(CEO)