

HORIBA

Medical

HORIBA India Private Limited

246, Okhla Industrial Estate Phase-III,
New Delhi 110020, India
Tel : +91 (11) 4646 5000 / 4669 5001
Fax : +91 (11) 4646 5020
http : //www.horiba.com
CIN : U73100DL2006PTC153232

HIN/MED/2023-2024 /10047895
20th JUNE 2023

CALIBRATION CERTIFICATE

This is to certify that the Hematology Analyzer Horiba YUMIZEN H550 bearing serial number 204YAXH03682 installed at KRSNAA DIAGNOSTIC CENTRE HOSHIARPUR as calibrated on 20TH JUNE 2023.

Calibrator : ABX MINOCAL
Lot No. : CX483
Expiry Date : 05th JULY 2023

The reports of Blank Cycle, Repeatability and Calibration Values were all found in acceptable range.

Next calibration cycle is due on 19TH JUNE 2024



Shrish Dixit

(Head- Products & Marketing)
For Horiba India Pvt. Ltd.

246, Okhla Industrial Estate Phase-
III, New Delhi 110020, India
Tel : -r91 (11) 4546 5000 / 46695001
Fax : +91 (11) 4646 5020
http : //www.horiba.com
CIN : U73100DL2006PTC1M232